

Instrument : Erba EM 200

| Analyte                  | Standard Unit | Result Value | Mean   | Z-Score | RMZ   |
|--------------------------|---------------|--------------|--------|---------|-------|
| ⓈⓈ Glucose               |               |              |        |         |       |
| ⓈⓈ Urea                  | mg/dl         | 74.00        | 83.27  | -1.13   | -1.13 |
| ⓈⓈ Creatinine            | mg/dl         | 24.56        | 24.74  | 0.11    | 0.29  |
| ⓈⓈ Total Protein         | mg/dl         | 1.20         | 0.95   | 0.48    | 3.05  |
| @ Albumin                | g/dL          | 8.00         | 7.06   | 0.86    | 1.42  |
| ⓈⓈ Uric Acid             | g/dL          | 3.70         | 4.18   | -1.22   | -1.14 |
| ⓈⓈ Calcium               | mg/dl         | 4.35         | 4.66   | -0.50   | -0.19 |
| ⓈⓈ Phosphorous Inorganic | mg/dl         | 7.89         | 8.53   | -0.96   | -0.44 |
| ⓈⓈ Cholesterol           | mg/dl         | 4.10         | 4.64   | -0.18   | 6.53  |
| ⓈⓈ Triglycerides         | mg/dl         | 172.00       | 154.80 | 1.26    | 0.85  |
| @ ALT                    | mg/dl         | 138.80       | 123.39 | 1.02    | 0.26  |
| @ AST                    | UL            | 5.00         | 5.27   | -0.58   | -0.13 |
| @ ALP                    | UL            | 16.00        | 13.52  | 0.89    | 0.57  |
|                          | UL            | 53.00        | 58.21  | -0.16   | 0.09  |

Instrument : Erba Chem 7

| Analyte            | Standard Unit | Result Value | Mean   | Z-Score | RMZ   |
|--------------------|---------------|--------------|--------|---------|-------|
| ⓈⓈ Bilirubin Total | mg/dl         | 0.20         | 0.20   | 0.67    | 0.34  |
| @ HDL              | mg/dl         | 40.69        | 33.77  | 1.13    | 1.07  |
| ⓈⓈ Amylase         | UL            | 59.00        | 54.55  | -0.45   | -0.33 |
| ⓈⓈ Chloride        | mmol/L        | 100.00       | 100.15 | -0.21   | -0.16 |
| ⓈⓈ Sodium          | mmol/L        | 133.10       | --     | -0.07   | 0.12  |
| ⓈⓈ Potassium       | mmol/L        | 5.08         | 5.25   | -0.21   | -0.43 |

Legend Ⓢ : Acceptable

| : 2.0 < z score < 3.0 - Warning

X : z score ≥ 3 - Unacceptable

# : Not Evaluated

Ⓢ : Delayed Result Entry

\* : Not considered for evaluation.

|                          |         |
|--------------------------|---------|
| Total Parameters         | 20      |
| Not Evaluated Parameters | 0       |
| Evaluated Parameters     | 20      |
| Outlier Parameters (X)   | 1       |
| EQAS Score Biochemistry  | 95.00 % |

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416,  
Commercial Building - 1A  
Kohinoor Mall, Kiroi Road, Kurla (W),  
Mumbai - 400070