

Instrument : ERBA H 360

Analyte	Standard Unit	Result Value	Mean	Z-Score	RM2
ⓐ Total Leucocytes (WBC) count	10 ⁹ /L	10330	9726.00	-1.3	0.01
ⓐ Erythrocyte (RBC) Count	mill/cu.mm	2.78	2.81	-0.08	-0.11
ⓐ Haemoglobin	g/dL	8.70	8.83	-0.51	-0.11
ⓐ MCV (Mean Corpuscular Volume)	fL	132.30	147.39	-1.31	-1.2
ⓐ MCH (Mean Corpuscular Hb)	pg	31.40	31.44	-0.21	0.1
ⓐ MCHC (Mean Corpuscular Hb Concentration)	g/dL	23.80	21.41	0.71	1.0
ⓐ Platelet Count	1000/ μ L	236.00	276.29	-1.45	-1.11
ⓐ Hematocrit	%	36.70	41.28	-1.01	-1.11
ⓐ RDW CV%	%	16.90	19.69	-1.46	-1.61

Legend ⓐ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score \geq 3 - Unacceptable

: Not Evaluated

ⓐ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	1
EQAS Score Haematology - CBC	88.89 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416,
Commercial Building - 1A
Kohinoor Mall, Kirod Road, Kurla (W),
Mumbai - 400070