

CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Page: 1 of 2

1 Name and Address of the Customer

**M/s.: Adichunchanagiri Hospital & Research Centre,
Balagangadharanatha Nagara - 571 448.
Nagamangala Taluk, Mandya District, Karnataka.**

2 Customer Reference

2.1 ULR No. : CC247324500059592F
2.2 SRF No. : 1292
2.3 Certificate No. : VI/24-25/1292-02
2.4 Format No. : VI-FRM-TH-006
2.5 Date of issue : 01-07-2024

3 Details Of Device Under Calibration(DUC).

3.1 Nomenclature : Refrigerator
3.2 Make / Model : Bluestar / VC375E
3.3 ID.No./SI.No. : ACL/RF-02 / 100000739397
3.4 Range : 2 to 8 °C
3.5 DUC Condition : Satisfactory
3.6 Calibration Procedure No. : SOP-38-06
3.7 No.of Pages : 2
3.8 Calibration Date : 29-06-2024
3.9 Calibration Due : 28-06-2025
3.10 Calibration done at : Onsite
3.11 Instrument Location : Central / Laboratory
3.12 Discipline : Thermal

4 Environmental Condition


Temperature 25.2 - 27.2 °C Humidity 50 - 60 %RH

5 Standards Used for Temperature Calibration

Sl. No.	Nomenclature	Make / Model	ID. No. / SL.No	Cal Agency	Certificate. No.	Validity
1	Paperless Recorder	Brain Child	208210012	VI, Bangalore	VI/24-25/INT-ET-411	15-10-2024
2	RTD Sensor	Frontier / ---	VI-RTD-01	VI Bangalore	VI/23-24/INT-TH-100-01	12-08-2024

6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI.
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL Symbol is as per NABL guidelines given in NABL-133

Calibrated By

Nived
(Calibration Engineer)

Checked By

Umesh D
(Onsite In-Charge)



CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No. **VII/24-25/1292-02** Page: **2 of 2**


Results:

Refrigerator:

Range: 2 to 8 °C

Sl. No.	Set Knob Position On DUC	STD Sensors Reading in (°C)
1	2	2.0
2	4	4.0
3	6	6.1
4	7	7.1

Pictorial Representation of Sensors Placement



Note :

1. S Indicates Placement of Standard Sensor

Conclusion / Remarks:

1. Measurement Uncertainty reported is at 95.45% confidence level with k = 2.

Calibrated By

Nived
Nived

(Calibration Engineer)

Checked By

Umesh D
Umesh D

(Onsite In-Charge)

