



## CERTIFICATE OF CALIBRATION

COMMITTED TO THE CUSTOMER SINCE - 1996



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

**1 Name and Address of the Customer** M/s.: **Adichunchanagiri Hospital & Research Centre,  
Balagangadharanatha Nagara - 571 448.  
Nagamangala Taluk, Mandya District, Karnataka.**

**2 Customer Reference**

2.1 ULR No. : CC247324500059593F  
2.2 SRF No. : 1292  
2.3 Certificate No. : **VI/24-25/1292-03**  
2.4 Format No. : VI-FRM-TH-006  
2.5 Date of issue : 01-07-2024

**3 Details Of Device Under Calibration(DUC).**

3.1 Nomenclature : **Temperature Indicator With Sensor of Deep Freezer**  
3.2 Make / Model : **Bluestar / CH5D300FHSW**  
3.3 ID.No./SI.No. : **ACL/DF-01 / 341-95106601**  
3.4 DUC Condition : **Satisfactory**  
3.5 Calibration Procedure No. : **SOP-38-06**  
3.6 No.of Pages : **2**  
3.7 Calibration Date : **29-06-2024**  
3.8 Calibration Due : **28-06-2025**  
3.9 Calibration done at : **Onsite**  
3.10 Instrument Location : **Central / Laboratory**  
3.11 Discipline : **Thermal**

**4 Environmental Condition**

Temperature 25.2 to 27.2 °C Humidity 50 to 60 %RH

**5 Standards Used for Temperature Calibration**

Sl. No.	Nomenclature	Make & Model	ID. No.	Certificate No / Cal Agency	Validity
1	Paperless Recorder	Brain Child / VR-18	VI/OS/PR-03	VI/24-25/INT-ET-422 / VI, Bangalore	09-10-2024
2	RTD Sensor	Frontier / ----	VI-RTD-01	VI/23-24/INT-TH-100-01 / VI, Bangalore	12-08-2024

**6 Note:**

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI.
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7. The usage of NABL Symbol is as per NABL guidelines given in NABL-133

Calibrated By  
*Nived*  
Nived  
(Calibration Engineer)

Checked By  
*F. Umesh D*  
Umesh D  
(Onsite In-Charge)



**CERTIFICATE OF CALIBRATION**



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.  
Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No. VI/24-25/1292-03


Page: 2 of 2

Results:

Range: -18 to -20 °C

Sl. No.	Set Knob Position On DUC	STD Sensors Reading in (°C)
1	2	-18.1
2	6	-20.3

Pictorial Representation of Sensors Placement



Note : 1. S1 Indicates Placement of Standard Sensor

Conclusion / Remarks:

1 Measurement Uncertainty reported is at 95.45% confidence level with k = 2.

Calibrated By

*Nived*  
Nived

(Calibration Engineer)

Checked By

*Umesh D*  
Umesh D

(Onsite In-Charge)

