



COMMITTED TO THE
CUSTOMER SINCE - 1996

VAIDYANATHESHWARA INSTRUMENTS PRIVATE LIMITED

CALIBRATION CERTIFICATE

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Contact : 080-23377266, Mob : 9448080177 / 9632221171 / 9964308118 | Email : viplgroup65@gmail.com Web : www.viplgroup.com

"CALIBRATION LABORATORY"

Date of issue : 01-07-2024

Sheet : 1 of 1

Format No. : VI-FRM-ME-RM-001		Certificate No.: VI/24-25/1292-07			
Customer Name and Address:		M/s.: Adichunchanagiri Hospital & Research Centre, Balagangadharanatha Nagara - 571 448. Nagamangala Taluk, Mandya District, Karnataka.			
SRF No.	1292	Received Condition		Satisfactory	
CALIBRATED INSTRUMENT / EQUIPMENT DETAILS					
Nomenclature	Rotary Mixer	Make	Techno Fab		
Range	0 to 30 RPM	Id. No.	ACL/RM-01		
Calibrated on	29-06-2024	Calibration due on	28-06-2025		
Calibration Done At	Onsite	Temperature / Humidity	25.2 - 27.2 °C / 50 - 60 % RH		
Instrument Location	Central Laboratory	Discipline	Mechanical		
Master Traceability Details					
SI. No.	Nomenclature	Make / Model	SI.No.	Certificate No. / Cal Agency	Validity
1	Digital Tachometer	Mextech / DT-2234C	S086943	VI/23-24/INT-ME-505 / VI, Bangalore	13-10-2024
The master equipments used are traceable to National Standards			Method of Calibration : Comparison		
Test Results & Test Details :					
SI. No.	Range (RPM)	DUC Set (RPM)	Standard Instruments Reading (RPM)	Error Observed (RPM)	
1	0 to 30 RPM	5	5.0	0.0	
2		10	10.0	0.0	
3		15	15.0	0.0	
4		20	20.1	0.1	
5		30	30.1	0.1	
Note :					
Determination of various readings of above Instrument are from respective reference standards by direct comparison method.					
Conclusion : Measured Reading are Reported.					
The Reported Results are valid only for the conditions of the received instruments at the time of and under the stated conditions of the calibration					

Calibrated By

Nived
Nived

(Calibration Engineer)

Checked By

F. Kanya
Umesh D

(Onsite In-Charge)

Authorized By

Gangadhar C.K
Gangadhar C.K

