

Instrument : Roche cobas c 311

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	33.00	33.21	0.16	0.28
@ Creatinine	mg/dl	0.90	1.00	-1.11	-0.42
@ Bilirubin Total	mg/dl	0.20	0.18	0.60	0.41
@ Bilirubin Direct	mg/dl	0.10	0.08	0.49	0.93
@ Total Protein	g/dL	7.20	7.10	0.36	0.24
@ Albumin	g/dL	4.40	4.23	1.11	0.16
@ Uric Acid	mg/dl	4.20	4.40	-0.55	-0.16
@ Calcium	mg/dl	8.90	8.82	0.28	0.30
@ Phosphorous Inorganic	mg/dl	4.10	4.18	0.07	0.30
@ Cholesterol	mg/dl	178.00	173.59	0.96	0.16
@ Triglycerides	mg/dl	154.00	150.40	0.42	0.56
@ HDL	mg/dl	23.00	29.40	-1.22	0.32
@ ALT	U/L	14.00	19.50	-0.84	0.81
@ AST	U/L	25.00	30.75	-0.63	0.47
@ ALP	U/L	81.00	81.65	-0.02	-0.11
@ GGT	U/L	24.00	23.81	0.17	0.40
@ Amylase	U/L	80.00	76.96	0.94	0.17
@ Chloride	mmol/L	106.00	98.94	0.57	0.28
@ Sodium	mmol/L	155.00	138.10	1.17	-0.20
@ Potassium	mmol/L	5.90	5.39	1.10	-0.38

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning

X : z score ≥ 3 - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	20
Not Evaluated Parameters	0
Evaluated Parameters	20
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirod Road, Kurla (W),
Mumbai - 400070