



Roche Diagnostics India Pvt. Ltd.
Field Service Report



Case number CAS-0019594722

Order Number WO-26808788 | Ser. Report Created Date 11/04/2024

Instrument Details

Customer details

Instrument/Module cobas c 311
Serial Number 1360-09

Customer Number 0052011871

Customer Name HIRA CLINICAL
LABORATORY

Street Address CIVIL HOSPITAL ROAD CIVIL
HOSPITAL

Zip 146001 - HOSHARPUR

Additional Details

Contact Name

Contact Phone

Service Activity Code PM Visit

Purpose of Visit

Preventive maintenance of machine

Performed Activities

Done Preventive Maintenance of machine as per checklist,Run samples machine working fine.

Time Report

Category	Start	End	Hours	Invoice Type
05913616001-Service labour time	11/04/2024 14:20	11/04/2024 15:39	1.30	Free of Charge

Travel

Category	Hours
Travel	0.02

Customer Parts Used

Part Number	Description	Quantity
05182522001-KIT MAINTENANCE 6 MONTHS COBAS C311	KIT MAINTENANCE 6 MONTHS COBAS C311	1
05182548001-KIT MAINTENANCE 1 YEAR COBAS C311	KIT MAINTENANCE 1 YEAR COBAS C311	1



 * HITACHI AUTOMATIC ANALYZER *

NAME
 S.NO.
 LOT

PCCC2
 C030134 001
 59539300

DATE 20/04/24 08:22:18
 OPERATOR ID bmserv

TEST	RESULT	UNIT	EXPECTED VALUE	ALARM
GLUCOSE	233.9	mg/dL	(208- 256)	
AST/SGOT	139.5	U/L	(124- 160)	
GGT	240	U/L	(208- 264)	
ALP	245	U/L	(220- 284)	
BIL-D	3.25	mg/dL	(2.50- 3.38)	
ALB	5.01	g/dL	(4.75- 6.03)	
ALT/SGPT	116.8	U/L	(110- 142)	
CREAT	3.46	mg/dL	(3.24- 4.12)	
UREA	130.2	mg/dL	(117- 141)	
CALCIUM	12.7	mg/dL	(11.8- 13.8)	
URIC ACID	9.6	mg/dL	(9.32- 11.28)	
TOTAL PROTEI	8.01	g/dL	(7.66- 8.98)	
CHOLESTEROL	175.4	mg/dL	(161- 193)	
TRIGLYCERIDE	219.5	mg/dL	(208- 252)	
PHOS2	7.86	mg/dL	(7.11- 8.75)	
BILT3	4.08	mg/dL	(3.21- 4.09)	
AMYL2	207	U/L	(173- 217)	
HDLc4	55.0	mg/dL	(45.2- 62.0)	
CRP4	49.0	mg/L	(44.7- 58.3)	

[Handwritten signature]