

Calibration Report

Sample ID CX485
Lot number CX485

Name ABX MINOCAL

Exp. date 05/09/2023
Modified on

Coefficient	WBC	RBC	HGB	HCT	PLT
New	1.091	1.006	0.907	0.904	1.165
Current	1.116	1.010	0.949	0.900	1.100
Target	9.10	4.53	13.0	36.1	270
Mean	9.31	4.55	13.6	36.1	255
CV(%)	2.41	1.19	0.71	1.26	3.23

Number of calibration run selected for coefficient calculation (minimum 5) 8/11

Sel.	Run Time	WBC (10 ³ /mm ³)	RBC (10 ⁶ /mm ³)	HGB (g/dL)	HCT (%)	PLT (10 ³ /mm ³)
✓	07/08/2023 11:33:54 AM	9.09	4.50	13.5 h	36.3	248 l
✓	07/08/2023 11:35:28 AM	9.20	4.59	13.5 h	36.8	254 l
✓	07/08/2023 11:36:43 AM	8.98	4.48	13.5 h	35.7	263
	07/08/2023 11:37:56 AM	9.27	4.63 h	13.6 h	36.7	274
✓	07/08/2023 11:39:57 AM	9.30	4.58	13.6 h	36.5	243 l
✓	07/08/2023 11:41:15 AM	9.44 h	4.56	13.7 h	36.3	251 l
✓	07/08/2023 11:42:26 AM	9.49 h	4.49	13.7 h	35.4	267
	07/08/2023 11:44:16 AM	9.55 h	4.65 h	13.8 h	36.9	257 l
	07/08/2023 11:45:25 AM	9.39 h	4.63 h	13.6 h	36.6	264
✓	07/08/2023 11:47:43 AM	9.30 h	4.54	13.6 h	36.0	257 l
✓	07/08/2023 11:49:01 AM	9.67 h	4.63 h	13.7 h	36.0	261