



SHIVA CALIBRATION LABORATORY

M.C.F-37, Gall No.-2 Navlu Colony, Railway Road, Near Metro Station Ballabgarh-121004 (Hr.)

Mobile : 7015724855, 7056221717, 9992348281

E-mail : shivacallab@gmail.com, sclcaltech@gmail.com, Website : www.shivalab.in



CC-3421

CALIBRATION CERTIFICATE

FORMAT NO.QS-22F

Page No.1 of 1

Name & Address of Client	Certificate /ULR No.	CC 34212210000299972F
M/S Careland Medical Systems 75-76, 1st floor SRC, Nilothi New Delhi-110041	Date of issue	12.09.2023
	Date of calibration	11/09/2023
	Calibration valid upto	10/09/2025
	Service request No & Date	2023/562/11.09.23
	UUC Receipt Date	11/09/2023

Equipment under calibration			
Name	Temperature Gauge meter	Least count	N.A
Make /Model No.	AMCO	calibration part	Sensor
Sr. No.	573945	Visual Inspection	Ok
ID. No.	CMS-5591	Zero Error	Nil
Range /Size	0 TO 300°C	Location	ON SITE

Standard Equipment used for calibration						
Sr. No.	Instrument Name	Make	Model/Sr.No	Certificate No.	Calibration Validity	Calibrated By
1	SPRT with Indicator (6.5 Digit Multimeter)	Thermo Electric/Fluke	21:12198:01 :01	CC33462200000983F	19/12/2023	Tycon Engineering

Environmental Condition	Temperature(Criteria)	(23±5)°C	Reference Standard	°C
	Temperature(Observed)	(25.5°C	Calibration Procedure	Temp.
	Relative Humidity (Criteria)	(40 to 60)%RH	Calibration PerformedAt	Site
	Relative Humidity (Observed)	51%RH	Discipline	temp .

Calibration Results


S.No	Standard	Actual value Displayed	Acceptable ± Uncertainty in percentage
1.	10°C	10°C	±0.5%
2.	50 °C	50 °C	±0.5%
3.	121°C	121°C	±0.5%

Remarks :--- Digital calibration

Note:

- This Certificate refers only the particular item submitted for calibration.
- This certificate shall not be reproduced except in full, without the written permission of the Chief Executive Shiva Calibration Laboratory , FARIDABAD.
- Results reported are valid at the time of and under the stated conditions of measurement..
- Laboratory Standards are traceable to National Standards.
- Calibration Certificates issued for Weight & Measurement parameters i.e Mass ,Balance, Volumetric equipment, Measuring Scale /Tapes etc. are for scientific purpose only should not be used for Trade / commercial use.

Calibrated by: 
 Calibration Engineer

 Approved by: 
 Technical Manager

End of the Report.



Careland Medical Systems

The Best Medical Solutions

Manufacturing of Pharmaceutical Refrigerators, Patient Care Equipments
& Service of All Type of Medical Equipments

GSTIN : 07BLZPS1938N1Z

Govt. Regd. No. 575

Email : careland.delhi@gmail.com

Mob.: +91-9953555905, +91-989911257

STANDARD CALIBRATION CERTIFICATE

Certificate no : DL24C021001CA	Certificate issued on : 21/03/2024
Date of calibration : 19/03/2024	Valid up to : 18/03/2025
Installed at : Laboratory	R.N.C date : 19/03/2025
Reference no : GEMC-511687717714137	

Equipment calibrated for : The Medical superintendent , Chacha Nehru Bal Chikitsalya
Govt. of NCT Delhi , Geeta Colony , New Delhi-31

Equipment under calibration

Equipment : Autoclave machine
Type : Digital

Make : Mac AVF 91(PROFAD)

Condition of equipment: Used in laboratory

Uncertainty
at 95%confidence level for a coverage factor k=2
(Single Position Comparison Method)
±0.85°C

Equipment used for calibrated by the device

Name	Temperature Gauge meter
Make /Model No.	AMCO
Sr. No.	573945
ID. No.	CMS-5591

Certificate /ULR No
CC 34212210000299972F
Valid up to 10/09/2025

S.no.	Temperature at UUC in (°C) (Single Position Comparison Method)	Avg. Standard Reading (°C)	Uncertainty at 95%confidence level for a coverage factor k= ±0.85 °C
1.	30	30	
2.	50	50	
3.	120	120	

This is certified that above mentioned equipment has been calibrated & tested as all over rating . The above equipment is within it's period of calibration which is traceable to the basis of National Standards. This certificate relates only to the item calibrated.

Calibrated by: Mr. Pramod singh
(Sr. Engineer)



Verified by



Authorized signatory
For CARELAND MEDICAL SYSTEMS





National Accreditation Board for
Testing and Calibration Laboratories

CERTIFICATE OF ACCREDITATION

SHIVA CALIBRATION LABORATORY

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2017

**"General Requirements for the Competence of Testing &
Calibration Laboratories"**

for its facilities at

MCF-37, GALI NO.2, NAVLU COLONY, RAILWAY ROAD, BALLABGARH, FARIDABAD, HARYANA, INDIA

in the field of

CALIBRATION

Certificate Number: CC-3421

Issue Date: 28/06/2022

Valid Until: 27/06/2024

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.
(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Name of Legal Identity : SHIVA CALIBRATION LABORATORY

Signed for and on behalf of NABL



N. Venkateswaran
Chief Executive Officer



SHIVA CALIBRATION LABORATORY

M.C.F-37, Gali No.-2 Navlu Colony, Railway Road, Near Metro Station Ballabgarh-121004 (Hr.)

Mobile : 7015724855, 7056221717, 9992348281

E-mail : shivacallab@gmail.com, sclcaltech@gmail.com, Website : www.shivalab.in



CC-3421

938N1
No. 57
rail.co
11257

CALIBRATION CERTIFICATE

FORMAT NO.QS-22F

Page No.1 of 1

Name & Address of Client	Certificate /ULR No.	CC 34212210000299972F
M/S Careland Medical Systems 75-76, 1st floor SRC, Nilothi New Delhi-110041	Date of issue	12.09.2023
	Date of calibration	11/09/2023
	Calibration valid upto	10/09/2025
	Service request No & Date	2023/562/11.09.23
	UUC Receipt Date	11/09/2023

Equipment under calibration

Name	Temperature Gauge meter	Least count	N.A
Make /Model No.	AMCO	calibration part	Sensor
Sr. No.	573945	Visual Inspection	Ok
ID. No.	CMS-5591	Zero Error	Nil
Range /Size	0 TO 300°C	Location	ON SITE

Standard Equipment used for calibration

Sr. No.	Instrument Name	Make	Model/Sr.No	Certificate No.	Calibration Validity	Calibrated By
1	SPRT with Indicator (6.5 Digit Multimeter)	Thermo Electric/Fluke	21:12198:01 :01	CC334622000000983F	19/12/2023	Tycon Engineering

Environmental Condition	Temperature(Criteria)	(23±5)°C	Reference Standard	°C
	Temperature(Observed)	(25.5°C	Calibration Procedure	Temp.
	Relative Humidity (Criteria)	(40 to 60)%RH	Calibration PerformedAt	Site
	Relative Humidity (Observed)	51%RH	Discipline	temp .

Calibration Results

S.No	Standard	Actual value Displayed	Acceptable ± Uncertainty in percentage
1.	10°C	10°C	±0.5%
2.	50 °C	50 °C	±0.5%
3.	121°C	121°C	±0.5%

Remarks :---- Digital calibration

Note:

- This Certificate refers only the particular item submitted for calibration.
- This certificate shall not be reproduced except in full, without the written permission of the Chief Executive Shiva Calibration Laboratory , FARIDABAD.
- Results reported are valid at the time of and under the stated conditions of measurement.
- Laboratory Standards are traceable to National Standards.
- Calibration Certificates issued for Weight & Measurement parameters i.e Mass ,Balance,Volumetric equipment, Measuring Scale /Tapes etc. are for scientific purpose only should not be used for Trade / commercial use.

Calibrated by: [Signature]
Calibration Engineer

Approved by: [Signature]
Technical Manager

End of the Report.



Careland Medical Systems

The Best Medical Solutions

Manufacturing of Pharmaceutical Refrigerators, Patient Care Equipments
& Service of All Type of Medical Equipments

GSTIN : 07BLZPS1938N1ZY

Govt. Regd. No. 5759

Email : careland.delhi@gmail.com

Mob.: +91-9953555905, +91-9899112579

STANDARD CALIBRATION CERTIFICATE

Certificate no : DL24C021001CA	Certificate issued on : 21/03/2024
Date of calibration : 19/03/2024	Valid up to : 18/03/2025
Installed at : Laboratory	R.N.C date : 19/03/2025
Reference no : GEMC-511687717714137	

Equipment calibrated for : The Medical superintendent , Chacha Nehru Bal Chikitsalya
Govt. of NCT Delhi , Geeta Colony , New Delhi-31

Equipment under calibration

Equipment : Autoclave machine Type : Digital	Make : Mac AVF 91(PROFAD)
Condition of equipment: Used in laboratory	Uncertainty at 95%confidence level for a coverage factor k=2 (Single Position Comparison Method) $\pm 0.85^{\circ}\text{C}$

Equipment used for calibrated by the device

Name	Temperature Gauge meter	Certificate /ULR No CC 34212210000299972F Valid up to 10/09/2025
Make /Model No.	AMCO	
Sr. No.	573945	
ID. No.	CMS-5591	

S.no.	Temperature at UUC in (°C) (Single Position Comparison Method)	Avg. Standard Reading (°C)	Uncertainty at 95%confidence level for a coverage factor k= $\pm 0.85^{\circ}\text{C}$
1.	30	30	
2.	50	50	
3.	120	120	

This is certified that above mentioned equipment has been calibrated & tested as all over rating . The above equipment is within it's period of calibration which is traceable to the basis of National Standards. This certificate relates only to the item calibrated.

Calibrated by: Mr. Pramod singh
(Sr. Engineer)



Authorized signatory
For CARELAND MEDICAL SYSTEMS

