



BHARATH INSTRUMENTS

OFFICE : #36/239, Angappa Naicken Street, Mannady, Chennai - 600 001.

NABL
ACCREDITED
LABORATORY

CALIBRATION CERTIFICATE

BI:CC.No:0425/2024

Name & Address : M/s. ICTC,
BJR CAR NICOBAR.

Master Instrument Reference : Tachometer, LT Make, DT-2236 Model,
Sl.No.:132076, Range: 5 to 99999 RPM, Calibrated at
PCCPL, Chennai - 600 008. Report No.:PC-12-22/ATL/5194-01,
Dtd: 05-12-2022. Valid upto 04-12-2024. (NABL)


Test Instrument Reference : CENTRIFUGE, REMI Make, Model: C-854/8,
Range: 0 to 3500RPM, Sl. No.: BI:0425/2024,
Identification No.: ICTC/BJR/CENT-01.

Ambient Temp. When tested : 30°C

TEST READINGS

Master Instrument Readings	Test Instrument Readings	Error (if any)	
		Allowed	Observed
rpm	rpm	± 1 %	rpm
503	500	"	-3
1015	1000	"	-15
2026	2000	"	-26
3042	3000	"	-42


Remarks: The above instruments well within the specified accuracy

Calibrated by : 
Calibrated on : 30-01-2024

Next Calibrn. Due on : 29-01-2025



For BHARATH INSTRUMENTS


Signatory