



# CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

## DEPARTMENT OF CLINICAL BIOCHEMISTRY

### CMC EXTERNAL QUALITY ASSURANCE SCHEME

#### MONTHLY SUMMARY REPORT - JUNE 2024

LAB NO: 2513

LAB NAME: **TIRUPATI DIAGNOSTIC CENTRE**

CONSTITUENT GROUP

**Chemistry I**

Date of Result Entered: 07/06/2024

Date of Result Published: 27/06/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	D.BILIRUBIN	Diazonium salt ( Colorimetric ) / Jendrassik	Roche c111, c311, Integra 400, c501, c502, c503, c702	131	1.29	10.64	0.14	0.98 mg/dL	-2.26	0.02

u\* - method of uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (Action Signal)

#### LAB ADDRESS

TIRUPATI DIAGNOSTIC CENTRE,  
RAMKRISHNA ROAD, J N ROY RD,  
KHOSHBAGAN,  
West Bengal

#### Coordinator Contact Details:

E-Mail: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential.  
CMC EQAS does not sub contract any components.  
\*\*\*\*\* End of Report \*\*\*\*\*

*Pamela Christudoss*

Dr. Pamela Christudoss,  
CMC EQAS Coordinator,  
Christian Medical College, Vellore.