



**PROFICIENCY TESTING REPORT**  
**ISHIM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029

*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 5579

Distribution No. : 163-N

Month/Year: April/2024

Instrument ID: HORIBA CARE

Model Name : ABXMICROS 60 CT

Serial No. : 6040T95919

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra ( Prof. & Head), Hematology, AIIMS, Delhi,  
 Tel: 9013085730, E-Mail : info@ishtmainseqap.com

Date of issue &amp; status of the report: 04-07-2024[Final].

**CBC and Retic Assessment**

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /pl	1	6.29	6.07	12.36	13.1	0.075	-0.40	0.22	0.1	0.006	1.25
RBC x10 <sup>6</sup> /pl	1	4.8	4.76	9.56	9.74	0.012	-0.61	0.04	0.04	0.003	0.00
Hb g/dl	1	12.1	12.1	24.2	23.7	0.022	0.84	0	0.1	0.006	-0.67
HCT%	1	42.8	42.6	85.4	78.8	0.213	1.25	0.2	0.4	0.026	-0.54
MCV-f	1	89.5	89.2	178.7	161.9	0.401	1.42	0.3	0.3	0.021	0.00
MCH-Pg	1	25.4	25.2	50.6	48.8	0.058	1.28	0.2	0.2	0.013	0.00
MCHC-g/dl	1	28.4	28.3	56.7	60.1	0.174	-0.64	0.1	0.3	0.018	-0.67
PLT x10 <sup>3</sup> /pl	1	254	236	490	473	4.433	0.12	18	6	0.394	1.80
Retic %	2	9.5	8.5	18	14.7	0.200	0.56	1	0.5	0.032	0.84

**P.S. Assessment**

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbc=0, Poly=75 L=22, E=02, Mono/Promono=01, B1=0 P.M.=0, Mye=0, Meta=0, Other=0
RBC Morphology	3	Poly: 70-79, Lympho: 15-21, Mono: 2-6, Eosino: 1-3, blast/Promyelo/Myelo/Meta: 0
Diagnosis	3	Predominantly: Normocytic/Normochromic; Moderate: Sickle cells, Poikilocytosis, Target cells, Mild: Anisocytosis, Polychromasia, tear drop cells
		Sickle Cell Disease

### COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Parameters	S.No.	Total participants covered in the current dist. 163--N	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
BC x10 <sup>3</sup> /pl	1	315	315	81.27	79.68	6.03	7.94	12.7	12.38
BC x10 <sup>6</sup> /pl	1	315	315	87.3	91.75	7.62	3.81	5.08	4.44
Hb g/dl	1	315	315	86.98	89.84	7.94	3.17	5.08	6.99
HCT%	1	315	314	87.26	90.13	6.69	3.5	6.05	6.37
MCV-fl	1	315	315	89.84	90.16	5.71	2.54	4.45	7.3
MCH-Pg	1	315	315	86.98	92.7	7.62	3.49	5.4	3.81
CHC-g/dl	1	315	315	93.97	87.3	4.13	2.86	1.9	9.84
lt. x10 <sup>3</sup> /pl	1	315	315	97.46	88.57	1.9	7.3	0.64	4.13
ticCount%	2	315	238	94.96	90.34	4.2	5.04	0.84	4.62
Assessment	3	315	220	Satisfactory :80.33%, Borderline Sat. :14.60%, Unsatisfactory :5.07%					

#### Comments:

- 1). **Among Lab (EQA) : Results acceptable.**
- 2). **Within Lab (QA) : Precision acceptable.**

**Note-1:** EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**QA (Internal Quality Assurance) :** Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA) = (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (QA) = (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data. Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3: Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (S<sub>s</sub>) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample (x̄-y) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme


**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaimseqap.com](http://www.ishtmaimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra ( Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----