



RELIABLE TECHNICAL SERVICES

(Division of Reliable Technocare Pvt. Ltd.)

"RELIABLE HOUSE" 497/2834-35, Sant Tukaram Nagar,
Pimpri, Pune - 411018. MH, India. Telefax : 020-27421170

Cell : 7774055755, 7774055855, 7774058855, 7774022900

Email : reliable1010@gmail.com/reliabletechnocare@gmail.com

Web. : www.reliable.world



CC-2927

CALIBRATION CERTIFICATE

1.CUSTOMER	:-	Page No.	:- 1 of 1
Dr. Kakade Pathology Lab		NABL Accreditation No.	:- CC-2927
Kolhapur Naka, Karad		Certificate No.	:- 24.04.16.001
Temperature. (°C)	:- 22.5	Date of Received	:- 16.04.2024
Relative Humidity (% RH)	:- 51	Date of Calibration	:- 16.04.2024
Location of calibration	:- In Lab/On site	Next Calibration Due On	:- 15.04.2025
Condition of Item	:- OK	Calibration method No.	:- RTS-WI-06
		Date of Issue	:- 18.04.2024
		ULR No.	:- CC292724000003801F

2. Description of Item

Name	:- Centrifuge	Range	:- 3500	RPM
Id No	:- GEL C-7811	Least Count	:- -	
Make	:- REMI			
Type	:- --	Location	:- LAB	

3.Details of Equipment used for calibration

Name	:- DIG. TACHOMETER
Certificate No.	:- HT/CC/230922-12/001
Certified By	:- HI TECH
ID/Sr. No.	:- RTS-RPM-03
Calibration Validity	:- 26.09.2024

Discipline :- Mechanical Calibration Group :- Acceleration, Speed & Acoustics-Speed

4.Calibration Results

Calibration Points	Standard Reading	UUC Reading	Error in	Expanded Uncertainty in±
RPM	RPM	RPM	RPM	%
3500	3506	3500	-6	0.13

Note:

- 1)The reported uncertainty is the expanded uncertainty in measurement obtained by multiplying the standard uncertainty by the coverage factor $k=2$, which corresponds to a coverage probability of approximately 95% for normal distribution
- 2) This certificate refers only to the particular item submitted for calibration. UUC stands for Unit Under Calibration.
- 3) The calibration results reported in the certificate are valid at the time of and under the stated conditions of measurement.
- 4) Calibration point were selected as per customer specifications.
- 5) This certificate shall not be reproduced, except in full unless written permission for the publication of an approved abstract has been obtained from the Technical Manager of "Reliable Technical Services, Pune".
- 6) The Standard used are traceable to National / International Standard.


Calibrated By

S.Kore

Calibration Engineer

Sagar Kore

RF-21, R0



Approved By

R.Sable

Technical Manager

RAHUL SABLE

