



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2024



PC-1024

Lab Name **ISHA DIAGNOSTICS & RESEARCH PVT LTD**Lab No **4247**Constituent Group **Chemistry I**Date of Result Entered : **17/06/2024**PT item **Lyophilized human serum based**Date of Report Published : **27/06/2024**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE	GOD-POD	Any Analyser (Automation / Semi Automation)	1290	303.92	5.91	17.97	334.06 mg/dL	1.68	1.00
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	1166	78.13	7.68	6.00	64.4 mg/dL	-2.29	0.35
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation)	957	1.75	10.48	0.18	1.7 mg/dL	-0.27	0.01
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrossik	Any Analyser (Automation / Semi Automation)	1161	2.46	18.18	0.45	3.08 mg/dL	1.39	0.03
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	1256	5.37	7.71	0.41	5.25 g/dL	-0.29	0.02
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	1226	3.24	8.36	0.27	3.52 g/dL	1.03	0.02
7	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	1205	4.11	15.24	0.63	4.41 mg/dL	0.48	0.04
8	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	1235	114.86	5.63	6.47	114.95 mg/dL	0.01	0.37
9	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	1212	101.55	8.87	9.01	85.11 mg/dL	-1.83	0.52
10	HDL	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	1000	26.27	11.71	3.08	26.64 mg/dL	0.12	0.19
11	SODIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1464	125.91	3.67	4.62	123 mmol/L	-0.63	0.24
12	POTASSIUM	ISE - Direct	Any Analyser (Automation / Semi Automation)	1475	3.11	6.92	0.22	2.92 mmol/L	-0.88	0.01
13	CHLORIDE	ISE - Direct	Any Analyser (Automation / Semi Automation)	1188	93.81	5.06	4.75	89.9 mmol/L	-0.82	0.28
14	AST	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	1154	77.15	10.14	7.82	84.32 U/L	0.92	0.46
15	ALT	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	1180	205.62	16.45	33.83	269.06 U/L	1.88	1.97

16	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	990	50.25	13.95	7.01	53.43 U/L	0.45	0.45
----	-----	--------------------	--	-----	-------	-------	------	--------------	------	------

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

ISHA DIAGNOSTICS & RESEARCH PVT LTD

NO.311, 1st FLOOR, SAMPIG ROAD, BETWEEN 15th AND 16th CROSS, MALLESHWARAM

BENGALURU

KARNATAKA560003

Pamela Christudoss

Coordinator Contact Details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2024



Lab Name	ISHA DIAGNOSTICS & RESEARCH PVT LTD	Lab No	4247
Constituent Group	HbA1c	Date of Result Entered :	17/06/2024
PT item	Lyophilized human whole blood based	Date of Report Published :	27/06/2024

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	223	5.21	8.98	0.47	4.81 %	-0.85	0.06

u* - Method of Uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (action Signal)

LAB ADDRESS :

ISHA DIAGNOSTICS & RESEARCH PVT LTD
NO.311, 1st FLOOR, SAMPIG ROAD, BETWEEN 15th AND 16th CROSS, MALLESHWARAM
BENGALURU
KARNATAKA560003

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****