



CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096. Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer : M/s. ADICHUNCHANAGIRI HOSPITAL & RESARCH CENTRE Balagangadaranath Nagar - 571 448. Nagamangala Taluk, Mandya District Karnataka.

2 Customer Reference

- 2.1 ULR No : CC247324100025783F
2.2 SRF No : 2890
2.3 Certificate No. : VI/24-25/2890-03
2.4 Format No. : VI-FRM-ME-073
2.5 Dc No / Dc Date : Ref Letter / 29-06-2024
2.6 Receipt Date : 01-07-2024
2.7 Date Of Issue : 02-07-2025

3 Details Of Device Under Calibration(DUC).

- 3.1 Nomenclature : Micro Pipette
3.2 Make : Dragon Lab
3.3 Range : 100 - 1000µl
3.4 SL No. : YEA1CAG0018055
3.5 DUC Condition : Satisfactory
3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2022 E
3.7 No.of Pages : 2
3.8 Calibration Date : 02-07-2024
3.9 Calibration Due : 01-07-2025
3.10 Calibration done at : VI Volumetric Lab
3.11 Discipline : Mechanical (Mass & volume)

4 Environmental Condition

Temperature 22.2 - 22.5 °C Humidity 58.4 - 58.8 %RH

5 Standards Used for calibration

Table with 8 columns: SI. No., Nomenclature, Make / Model, SL No / ID No, Range, Certificate. No., Traceability to, Validity. Contains 2 rows of calibration data.

6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
6.3. Corrections/erasing, invalidate the Calibration Certificate.
6.4. Calibration of the DUC are traceable to National standards/International Standards
6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
6.6. Results Reported are valid at the time of and under the stated conditions of measurements.
6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By

Lomthkumar K.E (Calibration Engineer)

Checked By

Ranjith Kumar P (Lab- In-Charge)

Authorized By

Gangadhar C.K





CERTIFICATE OF CALIBRATION



CC-2473



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096. Ph : 080-23377266, Mob : 9986586789 / 9632221171 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: VI/24-25/2890-03

Page 2 of 2

Range: 100 - 1000µl

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty ±
100 - 1000µl	µl	µl	µl
	100	100.12	0.3
	500	500.25	2.0
	1000	1000.31	5.8

Note:

1. Visual Inspection : Found Well.
2. Trippl Distilled water is used to Calibrate the MicroPipette.

Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

[Signature]

Lohithkumar K.E
(Calibration Engineer)

Checked By

[Signature]

Ranjith Kumar P.
(Lab-In-Charge)

Authorized By

[Signature]

Gangadhar C.K

