

BHARATH INSTRUMENTS

OFFICE : #36/239, Angappa Naicken Street, Mannady, Chennai - 600 001.

NABL
ACCREDITED
LABORATORY

CALIBRATION CERTIFICATE

BI:CC.No:0390/2024

Name & Address

: ICTC,
CHC RANGAT.

Master Instrument
Reference

: Tachometer, LT Make, DT-2236 Model,
Sl.No.:132076, Range: 5 to 99999 RPM, Calibrated at
PCCPL, Chennai – 600 008. Report No.:PC-12-22/ATL/5194-01,
Dtd: 05-12-2022. Valid upto 04-12-2024. (NABL)

Test Instrument
Reference

: CENTRIFUGE , REMI Make, Model: R-304,
Range: 0 to 7000RPM, Sl. No.: ZGLN-32705 (BI:0390/2024),
Identification No.: CHE/CENT/ICTC/CENTR-01.

Ambient Temp.When tested : 30°C

TEST READINGS

Master Instrument Readings	Test Instrument Readings	Error (if any)	
		Allowed	Observed
rpm	rpm	$\pm 1 \%$	rpm
504	500	"	-4
1010	1000	"	-10
3015	3000	"	-15
5029	5000	"	-29
6535	6500	"	-35

Remarks: The above instruments well within the specified accuracy

Calibrated by

: *C. V. Ravi*

Calibrated on

: 30-01-2024

Next Calibn. Due on : 29-01-2025



For BHARATH INSTRUMENTS

A. K. S.

Signatory