



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3231

Distribution No.: 163-H

Month/Year: March/2024

Instrument ID: Sysmex 5 part analyser

Model Name.: XN 350

Serial No.: 11487

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra ( Prof. & Head), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : info@ishtmaiimseqap.com

Date of issue &amp; status of the report: 24-05-2024[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	10.23	10.11	20.34	15.8	0.082	3.29	0.12	0.11	0.014	0.07
RBC x10 <sup>6</sup> /µl	1	4.68	4.57	9.25	9.15	0.013	0.40	0.11	0.04	0.004	1.57
Hb g/dl	1	12.4	12.3	24.7	24.2	0.030	0.84	0.1	0.1	0.010	0.00
HCT%	1	42.3	41.7	84	77.6	0.243	1.25	0.6	0.4	0.032	0.54
MCV-fl	1	91.2	90.4	181.6	168.7	0.425	1.39	0.8	0.2	0.026	2.02
MCH-Pg	1	27.1	26.3	53.4	52.8	0.083	0.34	0.8	0.2	0.022	2.70
MCHC-g/dl	1	29.7	29.1	58.8	62.2	0.191	-0.88	0.6	0.3	0.016	1.01
Plt. x10 <sup>3</sup> /µl	1	293	287	580	475	3.248	1.73	6	7	0.553	-0.15
Retic %	2	5.1	5	10.1	10.5	0.220	-0.08	0.1	0.5	0.034	-0.90

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=14 , Poly=60 L=34, E=01, Mono/Promono=05 , B1=0 P.M.=0, Mye=0, Meta=0, Other=Hypersegmented neutrophils seen. Cytoplasmic vacuolation seen	Poly: 60-68, Lympho: 24-33, Mono: 2-5, Eosino: 1-2, blast/Promyelo/Myelo/Meta: 0-5		
RBC Morphology	3	Predominantly: Microcytic hypochromic and Normocytic normochromic; Mild anisocytosis, few macrocytes seen: Moderate: poikilocytosis, Target cells, few schistocytes, occasional spherocytes	Predominantly: Normocytic/Normochromic, Microcytic, Hypochromic, Moderate: Anisopoikilocytosis, Target cells Mild: Elliptocytes		
Diagnosis	3	Dimorphic anemia with hemolytic component	Thalassemia		

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 163--H	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	167	165	80.61	89.7	5.45	3.64	13.94	6.66
<b>RBC x10<sup>6</sup>/µl</b>	1	167	167	89.22	89.22	4.19	4.79	6.59	5.99
<b>Hb g/dl</b>	1	167	167	89.82	91.02	4.19	4.19	5.99	4.79
<b>HCT%</b>	1	167	165	96.97	89.09	3.03	4.85	0	6.06
<b>MCV-fl</b>	1	167	165	97.58	96.36	2.42		0	3.64
<b>MCH-Pg</b>	1	167	165	91.52	91.52	3.64	4.24	4.84	4.24
<b>MCHC-g/dl</b>	1	167	165	93.33	89.7	3.64	3.64	3.03	6.66
<b>Plt. x10<sup>3</sup>/µl</b>	1	167	165	89.09	91.52	7.88	5.45	3.03	3.03
<b>ReticCount%</b>	2	167	135	96.3	93.33	3.7	5.19	0	1.48
<b>PS Assessment</b>	3	167	124	Satisfactory :90.43%, Borderline Sat. :2.39%, Unsatisfactory :7.18%					

**\*Comments:**

- 1). Among Lab (EQA) : CBC result for WBC unacceptable, may be due to random/human error**
- 2). Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to  $\pm 2$ : Acceptable, Z score  $\pm 2$  to  $\pm 3$  :Warning Signal, Z score  $> \pm 3$  : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to  $\pm 2$ " are texted in green colour. Z score value between " $\pm 2$  to  $\pm 3$ " are texted in orange colour. Z score value  $> \pm 3$  are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra ( Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----