

Patient Report **Result Reprint** Test Statistics Calibration Trace Calibration Monitor Error Log Reaction Curve Other

- Patient Entry(F2)
- Test Parameter(F3)
- Profiles / Calc(F4)
- QC/Calibration(F5)
- Consumables(F6)
- Status Monitor(F7)
- Search(F8)
- Master
- Utility(F11)
- Service Check
- Maintenance(F12)
- Settings
- Wake Up
- Shut Down
- Help (F1)

Last Batch Patients Controls Calibration

Print Lab details Mark 2SD/3SD

Deleted Test

Filter Test : ALL Consumable : NORM Batch : Total Tests : 14

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Result Date	Curve	Mean	SD	Interval (3SD)	
<input type="checkbox"/>	1	S022358	NORM	SGPTD	37.8	U/L	-1SD	07/20/2024 08:...	56500	40.53	2.7000	32.43 - 48.63
<input type="checkbox"/>	3	S022358	NORM	SGOTD	38.4	U/L		07/20/2024 08:...	56502	38.61	2.5700	30.9 - 46.32
<input type="checkbox"/>	5	S022358	NORM	ALPU	106	U/L	-1SD	07/20/2024 08:...	56504	117.25	7.8300	93.76 - 140.74
<input type="checkbox"/>	7	S022358	NORM	GGT	60.7	U/L		07/20/2024 08:...	56506	61.42	4.0900	49.15 - 73.69
<input type="checkbox"/>	9	S022358	NORM	BIDD	0.85	mg/dL	+3SD	07/20/2024 08:...	56508	0.64	0.0500	0.49 - 0.79
<input type="checkbox"/>	11	S022358	NORM	BIT	1.58	mg/dL	+1SD	07/20/2024 08:...	56510	1.36	0.1100	1.03 - 1.69
<input type="checkbox"/>	13	S022358	NORM	UREA	31.0	mg/dL	-1SD	07/20/2024 08:...	56512	34.31	1.7200	29.15 - 39.47
<input type="checkbox"/>	15	S022358	NORM	CRENZ	1.19	mg/dL	+1SD	07/20/2024 08:...	56514	1.1	0.0600	0.92 - 1.28
<input type="checkbox"/>	17	S022358	NORM	GLU	104.5	mg/dL	+2SD	07/20/2024 08:...	56516	92.72	4.6400	78.8 - 106.64
<input type="checkbox"/>	19	S022358	NORM	TRIG	119.1	mg/dL	+1SD	07/20/2024 08:...	56518	112.22	5.6100	95.39 - 129.05
<input type="checkbox"/>	21	S022358	NORM	CHOL	147	mg/dL		07/20/2024 08:...	56520	143.81	7.1900	122.24 - 165.38
<input type="checkbox"/>	23	S022358	NORM	UA	5.7	mg/dL		07/20/2024 08:...	56522	5.74	0.2900	4.87 - 6.61
<input type="checkbox"/>	25	S022358	NORM	PRO	5.47	g/dL		07/20/2024 08:...	56524	5.56	0.3700	4.45 - 6.67
<input type="checkbox"/>	27	S022358	NORM	BIDD	0.72	mg/dL	+1SD	07/20/2024 09:...	56533	0.64	0.0500	0.49 - 0.79

Indication : Select Consumable

- Patient Entry(F2)
- Test Parameter(F3)
- Profiles / Calc(F4)
- QC/Calibration(F5)
- Consumables(F6)
- Status Monitor(F7)
- Search(F8)
- Master
- Utility(F11)
- Service Check
- Maintenance(F12)
- Settings
- Wake Up
- Shut Down

Patient Report **Result Reprint** Test Statistics Calibration Trace Calibration Monitor Error Log Reaction Curve Other

Last Batch Patients Controls Calibration

Print Lab details Mark 2SD/3SD

Deleted Test

From : 20-Jul-2024 To : 20-Jul-2024

Filter Test : ALL Consumable : PATH Batch : Total Tests : 13

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Result Date	Curve	Mean	SD	Interval (3SD)
<input checked="" type="checkbox"/>	2	S022359	PATH	SGPTD	143.2	U/L	07/20/2024 08:...	56501	144.94	9.6600	115.96 - 173.92
<input type="checkbox"/>	4	S022359	PATH	SGOTD	142.4	U/L	07/20/2024 08:...	56503	144.95	9.6500	116 - 173.9
<input type="checkbox"/>	6	S022359	PATH	ALPU	410	U/L	-1SD 07/20/2024 08:...	56505	471.7	31.4400	377.38 - 566.02
<input type="checkbox"/>	8	S022359	PATH	GGT	149.1	U/L	07/20/2024 08:...	56507	154.39	10.2900	123.52 - 185.26
<input type="checkbox"/>	10	S022359	PATH	BIDD	2.07	mg/dL	07/20/2024 08:...	56509	1.97	0.1600	1.49 - 2.45
<input type="checkbox"/>	12	S022359	PATH	BIT	5.40	mg/dL	07/20/2024 08:...	56511	5.27	0.4300	3.98 - 6.56
<input type="checkbox"/>	14	S022359	PATH	UREA	98.6	mg/dL	07/20/2024 08:...	56513	95.65	4.7800	81.51 - 110.19
<input type="checkbox"/>	16	S022359	PATH	CRENZ	4.01	mg/dL	07/20/2024 08:...	56515	4.02	0.2200	3.36 - 4.68
<input type="checkbox"/>	18	S022359	PATH	GLU	270.1	mg/dL	07/20/2024 08:...	56517	272.82	13.6400	231.9 - 313.74
<input type="checkbox"/>	20	S022359	PATH	TRIG	195.8	mg/dL	-1SD 07/20/2024 08:...	56519	210.79	10.5300	179.2 - 242.38
<input type="checkbox"/>	22	S022359	PATH	CHOL	213	mg/dL	+2SD 07/20/2024 08:...	56521	183.9	12.0400	147.78 - 220.02
<input type="checkbox"/>	24	S022359	PATH	UA	11.8	mg/dL	07/20/2024 08:...	56523	11.99	0.5900	10.22 - 13.76
<input type="checkbox"/>	26	S022359	PATH	PRO	8.90	g/dL	07/20/2024 08:...	56525	8.43	0.5600	6.75 - 10.11

Patient Entry(F2)

Test Parameter(F3)

Profiles / Calc(F4)

QC/Calibration(F5)

Consumables(F6)

Status Monitor(F7)

Search(F8)

Master

Utility(F11)

Service Check

Maintenance(F12)

Settings

Hibernate

Shut Down

Help(F1)

Last Batch Patients Controls Calibration Print Lab details Mark 2SD/3SD
 Deleted Test

Filter
 Test: ALL Consumable: PATH Batch: Total Tests: 16

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Result Date	Curve	Mean	SD	Interval (3SD)
2	S122325	PATH	SGPTD	125.6	U/L	-1SD	07/21/2024 08...	56622	138.44	9.2300	110.75 - 166.13
4	S122325	PATH	SGOTD	138.2	U/L		07/21/2024 08...	56624	139.72	9.3100	111.79 - 167.65
6	S122325	PATH	ALPJI	390	U/L	-1SD	07/21/2024 08...	56626	439.39	29.2900	351.52 - 527.26
8	S122325	PATH	GGT	146.7	U/L	+1SD	07/21/2024 08...	56628	136.28	9.0900	109.01 - 163.55
10	S122325	PATH	BID0	2.09	ng/dL		07/21/2024 08...	56630	2.11	0.1800	1.57 - 2.65
12	S122325	PATH	BIT	5.28	ng/dL		07/21/2024 08...	56632	5.26	0.4400	3.94 - 6.58
14	S122325	PATH	UREA	100.7	mg/dL	+1SD	07/21/2024 08...	56634	91.6	4.5800	77.86 - 105.34
16	S122325	PATH	CRENZ	4.21	ng/dL	+1SD	07/21/2024 08...	56636	3.89	0.2200	3.23 - 4.55
18	S122325	PATH	GLU	270.0	mg/dL		07/21/2024 08...	56638	257.35	12.8700	218.74 - 295.96
20	S122325	PATH	TRIG	193.4	mg/dL	-1SD	07/21/2024 08...	56640	208.85	10.4400	177.53 - 240.17
22	S122325	PATH	CHOL	207	mg/dL		07/21/2024 08...	56642	216.08	10.8000	183.68 - 248.48
24	S122325	PATH	UA	11.5	mg/dL		07/21/2024 08...	56644	11.4	0.5700	9.69 - 13.11
26	S122325	PATH	CA	12.3	mg/dL	+3SD	07/21/2024 08...	56646	10.56	0.5300	8.97 - 12.15
28	S122325	PATH	ALBO	4.67	g/dL	F...	07/21/2024 08...	56648	4.95	0.2500	4.2 - 5.7
30	S122325	PATH	PRO	8.26	g/dL		07/21/2024 08...	56650	8.53	0.5700	6.82 - 10.24
31	S122325	PATH	CA	11.6	mg/dL	+1SD	07/21/2024 08...	56651	10.56	0.5300	8.97 - 12.15

Indication: Select Consumable

MultikL v2021.C03 [EM 200] TLS - Reports - Result Reprint

Patient Report
 Result Reprint
 Test Statistics
 Calibration Trace
 Calibration Monitor
 Error Log
 Reaction Curve
 Other

Last Batch
 Patients
 Controls
 Calibration

Print Lab details
 Mark 250/350

Deleted Test

Filter: Test: ALL
 Consumable: NORM
 Batch:
 Total Tests: 15

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Result Date	Curve	Mean	SD	Interval (3SD)
1	5122324	NORM	SGPTD	40.4	U/L	-1SD	07/21/2024 08:...	56621	43.3	2.8900	34.63 - 61.97
3	5122324	NORM	SGOTD	36.6	U/L	-1SD	07/21/2024 08:...	56623	39.39	2.6300	31.5 - 47.28
5	5122324	NORM	ALPU	99	U/L	-1SD	07/21/2024 08:...	56625	113.04	7.5400	90.42 - 135.66
7	5122324	NORM	GGT	60.7	U/L		07/21/2024 08:...	56627	62.88	4.1900	50.32 - 75.46
9	5122324	NORM	BIDD	0.81	mg/dL		07/21/2024 08:...	56629	0.86	0.0700	0.65 - 1.07
11	5122324	NORM	BIT	1.59	mg/dL		07/21/2024 08:...	56631	1.52	0.1300	1.13 - 1.91
13	5122324	NORM	UREA	32.0	mg/dL	+1SD	07/21/2024 08:...	56633	29.77	1.4900	25.3 - 34.24
15	5122324	NORM	CRENZ	1.29	mg/dL	+2SD	07/21/2024 08:...	56635	1.14	0.0600	0.96 - 1.32
17	5122324	NORM	GLU	104.5	mg/dL	+1SD	07/21/2024 08:...	56637	98.85	4.9400	84.03 - 113.67
19	5122324	NORM	TRIG	115.4	mg/dL		07/21/2024 08:...	56639	120.89	6.0400	102.77 - 139.01
21	5122324	NORM	CHOL	142	mg/dL		07/21/2024 08:...	56641	146.55	7.3300	124.56 - 168.54
23	5122324	NORM	UA	5.4	mg/dL		07/21/2024 08:...	56643	5.41	0.2700	4.6 - 6.22
25	5122324	NORM	CA	9.9	mg/dL	+2SD	07/21/2024 08:...	56645	8.75	0.4400	7.43 - 10.07
27	5122324	NORM	ALBD	3.30	g/dL	F ₁	07/21/2024 08:...	56647	3.55	0.1800	3.01 - 4.09
29	5122324	NORM	PRO	5.59	g/dL		07/21/2024 08:...	56649	5.85	0.3900	4.68 - 7.02

Date wise
 From : 19-Jul-2024
 To : 19-Jul-2024

(X) Deleted Test

Select All

Print

Invert Selection

Export

SHOW

Filter

Test

ALL

Consumable

NORM

Batch

Total Tests : 13

Sr #	Lot #	Consumable	Test	Result	Unit	Flag	Result Date	Curve	Mean	SD	Interval (3SD)	
<input checked="" type="checkbox"/>	1	S022358	NORM	SGPTD	34.3	U/L	-2SD	07/19/2024 08:...	56398	40.53	2.7000	32.43 - 48.63
<input type="checkbox"/>	3	S022358	NORM	SGOTD	36.0	U/L	-1SD	07/19/2024 08:...	56400	38.61	2.5700	30.9 - 46.32
<input type="checkbox"/>	5	S022358	NORM	ALPU	100	U/L	-2SD	07/19/2024 08:...	56402	117.25	7.8300	93.76 - 140.74
<input type="checkbox"/>	7	S022358	NORM	BIDD	0.73	mg/dL	+1SD	07/19/2024 08:...	56404	0.64	0.0500	0.49 - 0.79
<input type="checkbox"/>	9	S022358	NORM	BIT	1.24	mg/dL	-1SD	07/19/2024 08:...	56406	1.36	0.1100	1.03 - 1.69
<input type="checkbox"/>	11	S022358	NORM	UREA	29.4	mg/dL	-2SD	07/19/2024 08:...	56408	34.31	1.7200	29.15 - 39.47
<input type="checkbox"/>	13	S022358	NORM	CRENZ	1.10	mg/dL		07/19/2024 08:...	56410	1.1	0.0600	0.92 - 1.28
<input type="checkbox"/>	15	S022358	NORM	GLU	99.2	mg/dL	+1SD	07/19/2024 08:...	56412	92.72	4.6400	78.8 - 106.64
<input type="checkbox"/>	17	S022358	NORM	TRIG	107.5	mg/dL		07/19/2024 08:...	56414	112.22	5.6100	95.39 - 129.05
<input type="checkbox"/>	19	S022358	NORM	CHOL	133	mg/dL	-1SD	07/19/2024 08:...	56416	143.81	7.1900	122.24 - 165.38
<input type="checkbox"/>	21	S022358	NORM	UA	5.4	mg/dL	-1SD	07/19/2024 08:...	56418	5.74	0.2900	4.87 - 6.61
<input type="checkbox"/>	23	S022358	NORM	CA	9.0	mg/dL		07/19/2024 08:...	56420	8.97	0.4500	7.62 - 10.32
<input type="checkbox"/>	25	S022358	NORM	PRO	5.33	g/dL		07/19/2024 09:...	56422	5.56	0.3700	4.45 - 6.67