

**CALIBRATION REPORT**

To,  
Gautam Iabdhhi Diagnostic Center,  
Pune.

Date- 27/05/2024

As desired by you please find enclosed here with the report.

Name of System	<i>Microlyte Electrolytes Analyzer</i>
Sr. No.	EL19NV874
Calibration done on	27/05/2024
Next calibration Due on	26/05/2025
Calibrated By	Mr. Eknath Jadhav

**Instrument calibration Values :-**

Electrode	Na+	K+	Cl-
Results	132.8	4.2	97.1
Mean	139	3.96	97
Range	132-146	3.76-4.16	92.2-102
Cal. values	60.5	58.2	48.1

All value is within range.

Checked instrument all found ok.

With Regards



Crystal Diagnostivision  
Service Engineer.