

## **CALIBRATION CERTIFICATE**

This is to certify that Biochemistry Analyzer **TURBOSTAT PLUS** bearing Serial No.**ES1012204CN178**. Which has been installed at **ULTRA MODERN LABORATORY, COIMBATORE** has been tested on the following aspects

### 1. LIGHT SOURCE :

- \* Constant Light intensity for each filter.

### 2. CALIBRATION :

- \* The Absorbance Factor according to the Factory set calibration.
- \* The Read Cell Temperature according to the Factory set calibration.

### 3. FILTERS :

<b>FILTERS(NM)</b>	<b>AD</b>
340	50054
405	47961
450	49864
505	45921
546	42831
578	48524
630	50261

Quality of the individual filters viz.340, 405, 505, 545, 580 and 630 were checked for calibration through the instrument software and calibrated.

On complete testing of the unit we found that the unit is working satisfactorily and used to carry out the necessary functions.

Date of Calibration : 21/03/2024  
Next Calibration Due: 20/03/2025

For **CPC DIAGNOSTICS PVT LTD.,**

*B. Meyyan*

Authorised Signatory.



**CORPORATE OFFICE**

No.70/6,Old No.108/6,4<sup>th</sup> floor,Westminister,Dr.Radhakrishnan Salai,Mylapore,Chennai-600004.  
Phone No:044-23460168/169