

BHARATH INSTRUMENTS

OFFICE : #36/239, Angappa Naicken Street, Mannady, Chennai - 600 001.

NABL
ACCREDITED
LABORATORY

CALIBRATION CERTIFICATE

BI:CC.No:0402/2024

Name & Address

: ICTC,
PHC CAMPBELL BAY,

Master Instrument
Reference

: Tachometer, LT Make, DT-2236 Model,
Sl.No.:132076, Range: 5 to 99999 RPM, Calibrated at
PCCPL, Chennai – 600 008. Report No.:PC-12-22/ATL/5194-01,
Dtd: 05-12-2022. Valid upto 04-12-2024. (NABL)

Test Instrument
Reference

: CENTRIFUGE , REMI Make, Model: C-854/8,
Range: 0 to 3500RPM, Sl. No.: HDL10086 (BI:0402/2024),
Identification No.: PHC/CENT/ICTC/CENTR-01.

Ambient Temp.When tested : 30°C

TEST READINGS

Master Instrument Readings	Test Instrument Readings	Error (if any)	
		Allowed	Observed
rpm	rpm	± 1 %	rpm
503	500	"	-3
1011	1000	"	-11
2025	2000	"	-25
3033	3000	"	-33

Remarks: The above instruments well within the specified accuracy

Calibrated by

: *C. V. Rath*

Calibrated on

: 30-01-2024

Next Calibn. Due on : 29-01-2025



For BHARATH INSTRUMENTS

A. K. [Signature]

Signatory

LAB : #33, Jones Street, Mannady, Chennai - 600 001.
E-mail : bharath.electronics4u@yahoo.com / bharathinstruments.electronics@gmail.com
Mob : 98418 21719, 97908 50019, Tel : 044-4352 1719. GSTIN : 33AUVPK5991R1ZD
Dealing Sales, Service, Calibration