



# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-122	Page	: 1 of 1	
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: IS:12508:1988	
CALIBRATION DUE.	: 30 / 06 / 2025	C. P. No.	: ACS/CP/10	
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC34272400000936F Dt. 02.07.2024	Cal. At	: Site / Lab	
CALIBRATED FOR	: Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida.-201301 (UP)			
ITEM CALIBRATED	: RPM of Blood Roller Mixer			
	Make : Transasia	Range	: (30 to 35) RPM	
	L. Count : 1 RPM	ID No.	: DDPCRIH/Hem/Lab/BRM-01	
	Location : Ground Floor	Deptt.	: Hematology Lab	
ENVIRONMENT CONDITION:	Temp. { 30°C ± 10°C } R.H. {30% to 80%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1	08/09/2024
RESULTS	: Mechanical Discipline			
* U.C.INSTT. READS (rpm)	STANDARD READS (rpm)	ERROR (rpm)		
31	31.0	-		
33	33.0	-		
35	35.1	0.1		
	(average of readings)			

### REMARKS

: Expanded uncertainty of measurement at 95% confidence level is  $\pm 2.5$  rpm at coverage factor  $k=1.96$ .

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by



....End of Calibration Certificate....







# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 240701A/N-121	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 01 / 07 / 2024	<b>Ref. Std.</b>	: CG-11
<b>CALIBRATION DUE.</b>	: 30 / 06 / 2025	<b>C.P. No.</b>	: ACS/CP/07
<b>SRF No. &amp; Dt. OF RECPT</b>	: 240701A/N Dt. 01.07.2024	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC34272400000935F Dt. 02.07.2024	<b>Cal. At</b>	: Site / Lab
<b>CALIBRATED FOR</b>	: Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida.-201301 (UP)		
<b>ITEM CALIBRATED</b>	: Digital Temperature Indicator of Refrigerator		
	<b>Make</b> : Symbol	<b>Range</b> :	(-20 to 20) °C
	<b>L. Count</b> : 0.1°C	<b>ID No.</b> :	DDPCRRIH/Hem/Lab/RefDTI-01
	<b>Location</b> : Ground Floor	<b>Deptt.</b> :	Hematology Lab
<b>ENVIRONMENT CONDITION:</b>	Temp. { 30°C ± 10°C } R.H. { 30% to 80% }		
<b>STANDARD INSTRUMENTS USED :</b>			
<b>INSTT. NAME</b>	<b>ID No. / S. No.</b>	<b>TRACEABILITY</b>	<b>CERTIFICATE No.</b>
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPU/041123/1-1
			<b>DUE DATE</b>
			07/11/2024
<b>RESULTS</b>	: ET Discipline		
* U.C. INSTT. READS (°C)	<b>STANDARD READING (°C)</b>	<b>ERROR (°C)</b>	
2.1	2.0	0.1	
4.2	4.0	0.2	
6.3	6.0	0.3	
8.5	8.0	0.5	
(average of readings)			

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is  $\pm 1.00^\circ\text{C}$  at coverage Factor  $k=1.96$ .

**U.C.** : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmary statement not provided since not required.

Prepared by



....End of Calibration Certificate....







# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
 Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-120	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: CG-11
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/07
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000934F Dt. 02.07.2024	Cal. At	: Site / Lab
CALIBRATED FOR	: Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida.-201301 (UP)		
ITEM CALIBRATED	: Digital Temperature Controller of Hot Air Oven		
Make	: ACE	Range	: (0 to 300) °C
L. Count	: 1 °C	ID No.	: DDPRCRIH/HemLab/HAOV-01
Deptt.	: Hematology/Gr. Floor	S. No.	: 1822

ENVIRONMENT CONDITION: Temp. { 30°C ± 10°C } R.H. { 30% to 80% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/041123/1-1	07/11/2024

RESULTS		ET Discipline	
* U.C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)	
50	50.0	-	
101	100.0	1	
202	200.0	2	
300	300.0	-	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.61°C at coverage Factor k=1.96.

\*U.C.

: Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by



....End of Calibration Certificate....





# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
 Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-119	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/10
SRF No.& DLOF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000933F Dt. 02.07.2024	Cal. At	: Site / Lab

**CALIBRATED FOR** : Dr. DP Rastogi Central Research Institute of Homoeopathy  
 A-1/1, Sector-24, Noida.-201301 (UP)

<b>ITEM CALIBRATED</b>	: Centrifuge
Make	: Remi
L. Count	: 10 RPM
Deptt.	: Hematology Lab/Gr.Floor
Range	: (0 to 5250) RPM
ID No.	: DDPRCIH/HemLab/Cent-01
S. No.	: ZCEN-15396

**ENVIRONMENT CONDITION:** Temp. { 30°C ± 10°C } R.H. (30% to 80%)

**STANDARD INSTRUMENTS USED :**

INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1	08/09/2024

RESULTS : Mechanical Discipline			
* U. C. INSTT. READS (RPM)	Standard Measured (RPM)	Error (RPM)	
1000	1001.2	- 1.2	
2000	2002.3	- 2.3	
4000	4005.4	- 5.4	
5250	5255.7	- 5.7	
(average of readings)			

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 14 rpm of reading at coverage Factor k=1.96.

\*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by



....End of Calibration Certificate....







# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
 Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-118	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: CG-11
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/07
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000932F Dt. 02.07.2024	Cal. At	: Site / Lab

CALIBRATED FOR : Dr. DP Rastogi Central Research Institute of Homoeopathy  
 A-1/1, Sector-24, Noida.-201301 (UP)

ITEM CALIBRATED : Digital Temperature Indicator of Refrigerator  
 Make : Symbol Range : (-20 to 20) °C  
 L. Count : 0.1°C ID No. : DDPRCRIH/BioChemLab/RefDTI-04  
 Location : 1<sup>st</sup> Floor Deptt. : Biochemistry Lab

ENVIRONMENT CONDITION: Temp. { 30°C ± 10°C } R.H. {30% to 80%}

INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/041123/1-1	07/11/2024

* U.C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)
- 20.0	- 20.0	-
2.1	2.0	0.1
4.1	4.0	0.1
6.2	6.0	0.2
8.4	8.0	0.4
(average of readings)		

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

- \*U.C. : Under Calibration.
- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
  - The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
  - This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
  - The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
  - Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
  - Confirmity statement not provided since not required.

Prepared by



Auth. Signatory  
 S.K. Sharma  
 CEO/ITM  
 C-24, Sec-10, Noida

End of Calibration Certificate....





# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)  
C-23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 240701A/N-117	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 01 / 07 / 2024	<b>Ref. Std.</b>	: CG-11
<b>CALIBRATION DUE.</b>	: 30 / 06 / 2025	<b>C.P. No.</b>	: ACS/CP/07
<b>SRF No. &amp; Dt. OF RECPT</b>	: 240701A/N Dt. 01.07.2024	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC34272400000931F Dt. 02.07.2024	<b>Cal. At</b>	: Site / Lab
<b>CALIBRATED FOR</b>	: Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida-201301 (UP)		
<b>ITEM CALIBRATED</b>	: Digital Temperature Indicator of Refrigerator		
	Make : Symbol	Range	: (-20 to 20) °C
	L. Count : 0.1°C	ID No.	: DDPCRJIH/BioChem/Lab/RefDTI-03
	Location : 1 <sup>st</sup> Floor	Deptt.	: Biochemistry Lab
<b>ENVIRONMENT CONDITION:</b>	Temp. { 30°C ± 10°C} R.H. {30% to 80%}		

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/041123/1-1	07/11/2024

RESULTS :			
* U.C. INSTT. READS (°C)		STANDARD READING (°C)	ERROR (°C)
2.1		2.0	0.1
4.2		4.0	0.2
6.4		6.0	0.4
8.4		8.0	0.4
(average of readings)			

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is  $\pm 1.00^\circ\text{C}$  at coverage Factor  $k=1.96$ .

- \*U.C. : Under Calibration.
- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
  - The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
  - This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
  - The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
  - Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
  - Confirmity statement not provided since not required.

Prepared by



....End of Calibration Certificate....

Auth. Signatory  
S.K. Sharma  
CEO/ITM  
C-23, Sec-10, Noida





# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)  
'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-116	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/10
SRF No. & DLOF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000930F Dt. 02.07.2024	Cal. At	: Site / Lab

CALIBRATED FOR	: Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida.-201301 (UP)		
ITEM CALIBRATED	: RPM of Vortex Shaker		
Make	: N.A	Range	: (0 to 3000) rpm
L. Count	: 1rpm	ID No.	: DDPRCRIH/BioChemLab/V.Sh-01
Location	: 1 <sup>st</sup> Floor	Deptt.	: Biochemistry Lab

ENVIRONMENT CONDITION:	Temp. { 30°C ± 10°C}	R.H. {30% to 80%}
------------------------	----------------------	-------------------

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1	08/09/2024

RESULTS			
* U. C. INSTT. READS (RPM)	Mechanical Discipline	Standard Measured (RPM)	Error (RPM)
500		500.6	- 0.6
1000		1001.1	- 1.1
2000		2002.5	- 2.5
3000		3003.7	- 3.7
(average of readings)			

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 14 rpm of reading at coverage Factor k=1.96.

- \*U.C. : Under Calibration.
- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
  - The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
  - This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
  - The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
  - Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
  - Conformity statement not provided since not required.

Prepared By *[Signature]*



*[Signature]*  
Auth. Signatory  
S.K. Sharma  
CEO/TM  
C-23, Sec-10, Noida

....End of Calibration Certificate....





# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-115	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/10
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000929F Dt. 02.07.2024	Cal. At	: Site / Lab

CALIBRATED FOR : Dr. DP Rastogi Central Research Institute of Homoeopathy  
A-1/1, Sector-24, Noida.-201301 (UP)

ITEM CALIBRATED : RPM of Blood Roller Mixer			
Make	: SDS	Range	: (30 to 36) RPM
L. Count	: 1 RPM	ID No.	: DDPCRHIH/BioChemLab/BRM-01
Location	: 1 <sup>st</sup> Floor	Deptt.	: Biochemistry Lab

ENVIRONMENT CONDITION: Temp. { 30°C ± 10°C } R.H. {30% to 80%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1	08/09/2024

RESULTS : Mechanical Discipline			
* U.C. INSTT. READS (rpm)	STANDARD READS (rpm)	ERROR (rpm)	
30	30.1	0.1	
32	32.1	0.1	
34	34.2	0.2	
36	36.2	0.2	
(average of readings)			

REMARKS : Expanded uncertainty of measurement at 95% confidence level is ± 2.5 rpm at coverage factor k=1.96.

U.C. : Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use. Confirmation statement not provided since not required.

Prepared by 



....End of Calibration Certificate....







# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)  
'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 240701A/N-114	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 01 / 07 / 2024	<b>Ref. Std.</b>	: CG-11
<b>CALIBRATION DUE.</b>	: 30 / 06 / 2025	<b>C.P. No.</b>	: ACS/CP/07
<b>SRF No. &amp; Dt. OF RECPT</b>	: 240701A/N Dt. 01.07.2024	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC34272400000928F Dt. 02.07.2024	<b>Cal. At</b>	: Site / Lab
<b>CALIBRATED FOR</b>	: Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida.-201301 (UP)		

<b>ITEM CALIBRATED</b>	: Digital Temperature Indicator of Refrigerator		
<b>Make</b>	: Samsung	<b>Range</b>	: (-20 to 20) °C
<b>L. Count</b>	: 0.1°C	<b>ID No.</b>	: DDPRCRIH/BioChem/Lab/RefDTI-02
<b>Location</b>	: 1 <sup>st</sup> Floor	<b>Deptt.</b>	: Biochemistry Lab

**ENVIRONMENT CONDITION:** Temp. { 30°C + 10°C } R.H. {30% to 80%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/041123/1-1	07/11/2024

RESULTS :			
* U.C. INSTT. READS (°C)		STANDARD READING (°C)	ERROR (°C)
- 19.7		- 20.0	0.3
2.0		2.0	-
4.3		4.0	0.3
6.5		6.0	0.5
8.6		8.0	0.6
(average of readings)			

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is  $\pm 1.00^\circ\text{C}$  at coverage Factor  $k=1.96$ .

\*U.C. : Under Calibration.  
 Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.  
 The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.  
 This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.  
 The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.  
 Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.  
 Conformity statement not provided since not required.

Prepared by



....End of Calibration Certificate....

Auth. Signatory  
 S.K. Sharma  
 CEO/TM  
 C-23, Sec-10, Noida





# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)  
 'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
 Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 240701A/N-113	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 01 / 07 / 2024	<b>Ref. Std.</b>	: CG-11
<b>CALIBRATION DUE.</b>	: 30 / 06 / 2025	<b>C.P. No.</b>	: ACS/CP/07
<b>SRF No. &amp; Dt. OF RECPT</b>	: 240701A/N Dt. 01.07.2024	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC34272400000927F Dt. 02.07.2024	<b>Cal. At</b>	: Site / Lab

**CALIBRATED FOR** : Dr. DP Rastogi Central Research Institute of Homoeopathy  
 A-1/1, Sector-24, Noida.-201301 (UP)

<b>ITEM CALIBRATED</b>	: Digital Temperature Indicator of Refrigerator		
<b>Make</b>	: Symbol	<b>Range</b>	: (-20 to 20) °C
<b>L. Count</b>	: 0.1°C	<b>ID No.</b>	: DDPRCRIH/BioChem/Lab/RefDTI-01
<b>Location</b>	: 1 <sup>st</sup> Floor	<b>Deptt.</b>	: Biochemistry Lab

**ENVIRONMENT CONDITION:** Temp. { 30°C ± 10°C } R.H. {30% to 80%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/041123/1-1	07/11/2024

RESULTS	U.C. INSTT. READS (°C)	ET Discipline	STANDARD READING (°C)	ERROR (°C)
	2.1		2.0	0.1
	4.2		4.0	0.2
	6.2		6.0	0.2
	8.3		8.0	0.3
	(average of readings)			

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

**C.** : Under Calibration.  
 Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.  
 Calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.  
 Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.  
 Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.  
 Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.  
 Affirmity statement not provided since not required.

Prepared by



....End of Calibration Certificate....

Auth. Signatory  
 S.K. Sharma  
 CEO/ITM  
 C-23, Sec-10, Noida





# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)  
 C-23, SECTOR-10, NOIDA - 201301 (U.P.)  
 Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-112	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/10
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000926F Dt. 02.07.2024	Cal. At	: Site / Lab

CALIBRATED FOR : Dr. DP Rastogi Central Research Institute of Homoeopathy  
 A-1/1, Sector-24, Noida.-201301 (UP)

ITEM CALIBRATED	: Centrifuge
Make	: Remi
L. Count	: 10 RPM
Deptt.	: Biochemistry/1 <sup>st</sup> Floor
Range	: (0 to 5250) RPM
ID No.	: DDPRCIH/BioChemLab/Cent-02
S. No.	: ZDBN-04233

ENVIRONMENT CONDITION: Temp. { 30°C ± 10°C } R.H. { 30% to 80% }

INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1	08/09/2024

U. C. INSTT. READS (RPM)	Standard (RPM)	Measured (RPM)	Error (RPM)
1000	1001.2		- 1.2
2000	2002.4		- 2.4
4000	4005.4		- 5.4
5250	5256.4		- 6.4
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 14 rpm of reading at coverage Factor k=1.96.

U.C. : Under Calibration.  
 Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.  
 The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.  
 This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.  
 The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.  
 Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.  
 Conformity statement not provided since not required.

Prepared by



Auth. Signatory  
 S.K. Sharma  
 CEO/TM  
 C-23, Sec-10, Noida

.....End of Calibration Certificate.....





# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
 Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-111	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: IS:12508:1988
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/10
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000925F Dt. 02.07.2024	Cal. At	: Site / Lab

CALIBRATED FOR : Dr. DP Rastogi Central Research Institute of Homoeopathy  
 A-1/1, Sector-24, Noida.-201301 (UP)

ITEM CALIBRATED	: Centrifuge
Make	: Remi
L. Count	: 10 RPM
Deptt.	: Biochemistry/1 <sup>st</sup> Floor
Range	: (0 to 6000) RPM
ID No.	: DDPRCRIH/BioChemLab/Cent-01
S. No.	: ZFJN-34763

ENVIRONMENT CONDITION: Temp. { 30°C ± 10°C } R.H. { 30% to 80% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Tachometer	ACS/DTM/01/ 211-0155V	Godrej	G-230908-1-1	08/09/2024

RESULTS			
* U. C. INSTT. READS (RPM)	Mechanical Discipline	Standard Measured (RPM)	Error (RPM)
1000		1001.3	- 1.3
2000		2002.6	- 2.6
4000		4004.8	- 4.8
6000		6007.2	- 7.2
(average of readings)			


REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 14 rpm of reading at coverage Factor k=1.96.

C. : Under Calibration.  
 Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.  
 Calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.  
 Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.  
 Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.  
 Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.  
 Conformity statement not provided since not required.

Prepared by



....End of Calibration Certificate....







# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-110	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: ISO:8655:6
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/08
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000924F Dt. 02.07.2024	Cal. At	: Site / Lab

CALIBRATED FOR : Dr. DP Rastogi Central Research Institute of Homoeopathy  
A-1/1, Sector-24, Noida-201301 (UP)

ITEM CALIBRATED	: Micro Pipette	Range	: (100 to 1000) $\mu$ l
Make	: Sartorius	ID No.	: 14507075
L. Count	: 5 $\mu$ l	Deptt.	: Biochemistry Lab
Location	: 1 <sup>st</sup> Floor		

ENVIRONMENT CONDITION: Temp. { 25°C  $\pm$  4°C } R.H. {30% to 75%}

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

RESULTS :				
Mechanical Discipline				
Nominal Capacity ( $\mu$ l)	Graduation ( $\mu$ l)	Nominal Value ( $\mu$ l)	Volume at 20 °C ( $\mu$ l)	Error ( $\mu$ l)
1000	5	100	99.75	0.25
		500	500.28	(-) 0.28
		1000	1000.29	(-) 0.29

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is  $\pm$  2.2  $\mu$ l at coverage factor  $k=1.96$ .

U.C. : Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.

The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.

This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.

The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.

Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.

Confirmation statement not provided since not required.

Prepared by



....End of Calibration Certificate....







# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
 Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-109	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: ISO:8655:6
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/08
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000923F Dt. 02.07.2024	Cal. At	: Site / Lab
CALIBRATED FOR	: Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida-201301 (UP)		
ITEM CALIBRATED	: Micro Pipette		
	Make : Sartorius	Range	: (100 to 1000) $\mu$ l
	L. Count : 5 $\mu$ l	ID No.	: 14507070
	Location : 1 <sup>st</sup> Floor	Deptt.	: Biochemistry Lab
ENVIRONMENT CONDITION:	Temp. { 25°C $\pm$ 4°C } R.H. {30% to 75%}		

### STANDARD INSTRUMENTS USED :

INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

### RESULTS :

Nominal Capacity ( $\mu$ l)	Mechanical Discipline			
	Graduation ( $\mu$ l)	Nominal Value ( $\mu$ l)	Volume at 20 °C ( $\mu$ l)	Error ( $\mu$ l)
1000	5	100	99.86	0.14
		500	500.28	(-) 0.28
		1000	1000.28	(-) 0.28

### REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is  $\pm 2.2 \mu$ l at coverage factor  $k=1.96$ .

U.C. : Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law. Instrument has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use. Confirmation statement not provided since not required.

Prepared by *B1*



....End of Calibration Certificate....

Auth. Signatory  
 S.K. Sharma  
 CEO/ITM  
 C-23, Sec-10, Noida





# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
 Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-108	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: ISO:8655:6
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/08
SRF No.& Dt.OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC342724000000922F Dt. 02.07.2024	Cal. At	: Site / Lab
CALIBRATED FOR	: Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida-201301 (UP)		
ITEM CALIBRATED	: Micro Pipette		
Make	: Sartorius	Range	: (20 to 200) µl
L. Count	: 1 µl	ID No.	: 14602776
Location	: 1 <sup>st</sup> Floor	Deptt.	: Biochemistry Lab
ENVIRONMENT CONDITION:	Temp. { 25°C ± 4°C } R.H. {30% to 75%}		

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

RESULTS :				
Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
		20	19.61	0.39
200	1	100	99.74	0.26
		200	199.48	0.52

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 2.2 µl at coverage factor k=1.96.

C. : Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services. This Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law. Instrument has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use. Confirmation statement not provided since not required.

Prepared by



....End of Calibration Certificate....







# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
 Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-107	Page	: 1 of 1	
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: ISO:8655:6	
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/08	
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC34272400000921F Dt. 02.07.2024	Cal. At	: Site / Lab	
CALIBRATED FOR	: Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida.-201301 (UP)			
ITEM CALIBRATED	: Micro Pipette			
	Make : Sartorius	Range	: (20 to 200) $\mu$ l	
	L. Count : 1 $\mu$ l	ID No.	: 14602777	
	Location : 1 <sup>st</sup> Floor	Deptt.	: Biochemistry Lab	
ENVIRONMENT CONDITION:	Temp. { 25°C $\pm$ 4°C } R.H. {30% to 75%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

RESULTS :				
Mechanical Discipline				
Nominal Capacity ( $\mu$ l)	Graduation ( $\mu$ l)	Nominal Value ( $\mu$ l)	Volume at 20 °C ( $\mu$ l)	Error ( $\mu$ l)
200	1	20	19.85	0.15
		100	99.78	0.22
		200	199.71	0.29

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is  $\pm 2.2 \mu$ l at coverage factor  $k=1.96$ .

\*U.C. : Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.  
 The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.  
 This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.  
 The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.  
 Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.  
 Conformity statement not provided since not required.

Prepared by



....End of Calibration Certificate....







# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 240701A/N-106	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 01 / 07 / 2024	<b>Ref. Std.</b>	: ISO:8655:6
<b>CALIBRATION DUE.</b>	: 30 / 06 / 2025	<b>C.P. No.</b>	: ACS/CP/08
<b>SRF No. &amp; DL OF RECPT</b>	: 240701A/N Dt. 01.07.2024	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC34272400000920F Dt. 02.07.2024	<b>Cal. At</b>	: Site / Lab

**CALIBRATED FOR** : Dr. DP Rastogi Central Research Institute of Homoeopathy  
A-1/1, Sector-24, Noida.-201301 (UP)

<b>ITEM CALIBRATED</b>	: Micro Pipette		
<b>Make</b>	: Sartorius	<b>Range</b>	: (10 to 100) µl
<b>L. Count</b>	: 1 µl	<b>ID No.</b>	: 13655957
<b>Location</b>	: 1 <sup>st</sup> Floor	<b>Deptt.</b>	: Biochemistry Lab

**ENVIRONMENT CONDITION:** Temp. { 25°C ± 4°C } R.H. { 30% to 75% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

Nominal Capacity (µl)	Mechanical Discipline			Volume at 20 °C (µl)	Error (µl)
	Graduation (µl)	Nominal Value (µl)			
100	1	10		9.48	0.52
		50		49.50	0.50
		100		99.74	0.26

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=1.96.

I.C. : Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services. This Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law. Instrument has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use. No conformity statement not provided since not required.

Prepared by



Auth. Signatory  
S.K. Sharma  
CEO/IM  
C-23, Sec-10, Noida

....End of Calibration Certificate....





# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-105	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: ISO:8655:6
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/08
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000919F Dt. 02.07.2024	Cal. At	: Site / Lab

CALIBRATED FOR : Dr. DP Rastogi Central Research Institute of Homoeopathy  
A-1/1, Sector-24, Noida.-201301 (UP)

ITEM CALIBRATED		: Micro Pipette	
Make	: Sartorius	Range	: (10 to 100) µl
L. Count	: 1 µl	ID No.	: 6121043
Location	: 1 <sup>st</sup> Floor	Deptt.	: Biochemistry Lab

ENVIRONMENT CONDITION: Temp. { 25°C ± 4°C } R.H. { 30% to 75% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/D449927064	BlueBox	BBT/691/JAN/24	04/01/2025

RESULTS :				
Mechanical Discipline				
Nominal Capacity (µl)	Graduation (µl)	Nominal Value (µl)	Volume at 20 °C (µl)	Error (µl)
100	1	10	9.51	0.49
		50	49.75	0.25
		100	99.84	0.16

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 0.4 µl at coverage factor k=1.96.

U.C. : Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law. Instrument has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use. Conformity statement not provided since not required.

Prepared by



Auth. Signatory  
S.K. Sharma  
GEO/ITM  
C-23, Sec-10, Noida

...End of Calibration Certificate...





# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

<b>CERTIFICATE NO.</b>	: 240701A/N-104	<b>Page</b>	: 1 of 1
<b>CALIBRATED ON.</b>	: 01 / 07 / 2024	<b>Ref. Std.</b>	: DKD R-5-7
<b>CALIBRATION DUE.</b>	: 30 / 06 / 2025	<b>C.P. No.</b>	: ACS/CP/11
<b>SRF No. &amp; Dt. OF RCPT</b>	: 240701A/N Dt. 01.07.2024	<b>Phy. Inspection</b>	: O.K.
<b>ULR No. &amp; ISSUE DATE</b>	: CC34272400000918F Dt. 02.07.2024	<b>Cal. At</b>	: Site / Lab
<b>CALIBRATED FOR</b>	: Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida.-201301 (UP)		
<b>ITEM CALIBRATED</b>	: Digital Thermo Hygrometer		
	Make : HTC	Range : (0 to 50)°C, (10-99)%	
	L. Count : 0.1 °C, 1%RH	ID No. : DDPRCRIH/HemLab/D.Thgm-02	
	Location : Ground Floor	Deptt. : Hematology Lab	
<b>ENVIRONMENT CONDITION:</b>	Temp. { 25°C ± 4°C } R.H. {30% to 75%}		

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temperature Sensor with Indicator	ACS/RHTIS/01/20365872	DVG	CC352323000001002F	14/07/2024

RESULTS :					
* U. C. INSTT. READS					
		Thermal Discipline	STANDARD READING	ERROR	UNC(+)
(°C)	15.1		14.92	0.18	2.1
	25.4		24.94	0.46	2.1
	40.5		39.97	0.53	2.1
	(average of readings)				
(%)	29		29.95	-0.95	2.5
	59		59.93	-0.93	2.5
	91		89.99	1.01	2.5
	(average of readings)				

**REMARKS** : The Expanded uncertainty of measurement at 95% confidence level is at coverage Factor k=1.96.

**Remarks:** Under Calibration.  
 Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.  
 The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.  
 This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.  
 This Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.  
 This Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.  
 Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.  
 No firmity statement not provided since not required.

Prepared by



.....End of Calibration Certificate.....







# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)  
C-23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-103	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: DKD R-5-7
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/11
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000917F Dt. 02.07.2024	Cal. At	: Site / Lab

CALIBRATED FOR : Dr. DP Rastogi Central Research Institute of Homoeopathy  
A-1/1, Sector-24, Noida.-201301 (UP)

ITEM CALIBRATED	: Digital Thermo Hygrometer
Make	: HTC
Range	: (0 to 50)°C, (10-99)%
L. Count	: 0.1 °C, 1%RH
ID No.	: DDPRCRIH/HemLab/D.Thgm-01
Location	: Ground Floor
Deptt.	: Hematology Lab

ENVIRONMENT CONDITION: Temp. { 25°C ± 4°C } R.H. { 30% to 75% }

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temperature Sensor with Indicator	ACS/RHTIS/01/20365872	DVG	CC352323000001002F	14/07/2024

RESULTS : Thermal Discipline				
* U. C. INSTT. READS	STANDARD READING		ERROR	UNC(±)
(°C)	15.5	14.92	0.58	2.1
	25.4	24.85	0.55	2.1
	40.7	39.97	0.73	2.1
(average of readings)				
(%)	32	29.94	2.06	2.5
	63	59.94	3.05	2.5
	94	89.96	4.04	2.5
(average of readings)				

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is at coverage Factor k=1.96.

U.C. : Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services. This Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use. Conformity statement not provided since not required.

Prepared by



....End of Calibration Certificate....







# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)  
Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO.	: 240701A/N-102	Page	: 1 of 1
CALIBRATED ON.	: 01 / 07 / 2024	Ref. Std.	: DKD R-5-7
CALIBRATION DUE.	: 30 / 06 / 2025	C.P. No.	: ACS/CP/11
SRF No. & Dt. OF RECPT	: 240701A/N Dt. 01.07.2024	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272400000916F Dt. 02.07.2024	Cal. At	: Site / Lab

CALIBRATED FOR	: Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida.-201301 (UP)
----------------	----------------------------------------------------------------------------------------------------

ITEM CALIBRATED	: Digital Thermo Hygrometer		
Make	: HTC	Range	: (0 to 50)°C, (10-99)%
L. Count	: 0.1 °C, 1%RH	ID No.	: DDPRCRIH/BioChemLab/D.Thgm-02
Location	: 1 <sup>st</sup> Floor	Deptt.	: Biochemistry Lab

ENVIRONMENT CONDITION:	Temp. { 25°C ± 4°C}	R.H. {30% to 75%}
------------------------	---------------------	-------------------

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temperature Sensor with Indicator	ACS/RHTIS/01/20365872	DVG	CC352323000001002F	14/07/2024

RESULTS :					
* U. C. INSTT. READS					
		Thermal Discipline	STANDARD READING	ERROR	UNC(±)
(°C)	15.2		14.90	0.30	2.1
	25.5		24.89	0.61	2.1
	40.8		39.98	0.82	2.1
	(average of readings)				
(%)	31		29.98	1.02	2.5
	62		59.93	2.07	2.5
	97		89.98	7.02	2.5
	(average of readings)				

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is at coverage Factor  $k=1.96$ .

\*U.C. : Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above. The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above. This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services. The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use. Conformity statement not provided since not required.

Prepared by



....End of Calibration Certificate....







# ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

C-23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 9868309579 • E-mail : advance.cal@gmail.com



## Calibration Certificate

CERTIFICATE NO. : 240701A/N-101 Page : 1 of 1

CALIBRATED ON. : 01 / 07 / 2024 Ref. Std. : DKD R-5-7

CALIBRATION DUE. : 30 / 06 / 2025 C.P. No. : ACS/CP/11

SRF No. & Dt. OF RECPT : 240701A/N Dt. 01.07.2024 Phy. Inspection : O.K.

ULR No. & ISSUE DATE : CC34272400000915F Dt. 02.07.2024 Cal. At : Site / Lab

CALIBRATED FOR : Dr. DP Rastogi Central Research Institute of Homoeopathy A-1/1, Sector-24, Noida-201301 (UP)

ITEM CALIBRATED : Digital Thermo Hygrometer  
Make : HTC Range : (0 to 50)°C, (10-99)%  
L. Count : 0.1 °C, 1%RH ID No. : DDPRCRIH/BioChemLab/D.Thgm-01  
Location : 1<sup>st</sup> Floor Deptt. : Biochemistry Lab

ENVIRONMENT CONDITION: Temp. {25°C + 4°C} R.H. {30% to 75%}

STANDARD INSTRUMENTS USED :

INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
RH & Temperature Sensor with Indicator	ACS/RHTIS/01/20365872	DVG	CC352323000001002F	14/07/2024

RESULTS : Thermal Discipline

* U. C. INSTT. READS	STANDARD READING	ERROR	UNC(±)
(°C)			
15.2	14.93	0.27	2.1
25.5	24.87	0.63	2.1
40.7	39.98	0.72	2.1
(average of readings)			
(%)			
32	29.94	2.06	2.5
63	59.94	3.06	2.5
93	69.92	3.08	2.5
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is at coverage Factor  $k=1.96$ .

U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by



....End of Calibration Certificate....