

NABL  
ACCREDITED  
LABORATORY



# BHARATH INSTRUMENTS

OFFICE : #36/239, Angappa Naicken Street, Mannady, Chennai - 600 001.



## CALIBRATION CERTIFICATE

BI:CC.No:0399/2024

Name & Address : ICTC,  
PHC NANCOWRY.

Master Instrument Reference : Tachometer, LT Make, DT-2236 Model,  
Sl.No.:132076, Range: 5 to 99999 RPM, Calibrated at  
PCCPL, Chennai - 600 008. Report No.:PC-12-22/ATL/5194-01,  
Dtd: 05-12-2022. Valid upto 04-12-2024. (NABL)

Test Instrument Reference : CENTRIFUGE , REMI Make, Model: C-854/8,  
Range: 0 to 3500RPM, Sl. No.: HDCL-10100 (BI:0399/2024),  
Identification No.: PHC/CENT/ICTC/CENTR-01.

Ambient Temp.When tested : 30°C

### TEST READINGS

Master Instrument Readings	Test Instrument Readings	Error (if any)	
		Allowed	Observed
rpm	rpm	± 1 %	rpm
504	500	"	-4
1010	1000	"	-10
2022	2000	"	-22
3035	3000	"	-35

Remarks: The above instruments well within the specified accuracy

Calibrated by : *e.vath*

Calibrated on : 30-01-2024

Next Calibrn. Due on : 29-01-2025



For BHARATH INSTRUMENTS

*A.K.*

Signatory

LAB : #33, Jones Street, Mannady, Chennai - 600 001.  
E-mail : bharath.electronics4u@yahoo.com / bharathinstruments.electronics@gmail.com  
Mob : 98418 21719, 97908 50019, Tel : 044-4352 1719. GSTIN : 33AUVPK5991R1ZD

Dealing Sales Service Calibration