

CERTIFICATE OF CALIBRATION,  
QUALITY CONTROL AND MAINTENANCE

To,  
Isha Diagnostic Laboratory.  
Banglore, Karnataka.


This is to certify that the instrument Eurocount 5L Serial No. 862406062IEJCP is under Calibration. Quality control check up of the instrument has been the responsibility of MEDSOURCE OZONE BIOMEDICALS PVT. LTD.

Date of Calibration : 16/09/2023.

Calibration & quality control checks as recommended by the manufacturers in operator manual are being performed regularly and satisfactorily.

This calibration certificate is valid up to:- 15/09/2024.

For Medsource Ozone Biomedicals Pvt. Ltd.



Sherin Joseph  
National Service Manager  
Medsource Ozone Biomedicals Pvt. Ltd.



**Medsource Ozone Biomedicals Pvt. Ltd.**

Head Office: Plot No. 109, HSIIDC, Sector-31, Faridabad-121003, (Haryana), INDIA  
Registered Office: 238, Sant Nagar, East of Kailash, New Delhi-110065, INDIA  
Telephone: +91-129-4286600 | E-mail: mail@ozonebio.com | Website: www.ozonebio.com  
CIN:U85100DL2003PTC120213

Lot No.: CBCP060923

Mode: Closed-WB

File No.: 2

Expiry date: 21-11-2023

Level: M

Date	17-11-2023	17-11-2023	17-11-2023
Time	13:33	13:34	13:37
WBC	7.32	7.29	7.41
NEU%	59.6	60.9	60.5
LYM%	33.5	34.4	33.7
MON%	2.6	2.4	3
EOS%	2.9	1.6	2.5
BAS%	1.4	0.7	0.3
NEU#	4.37	4.43	4.48
LYM#	2.45	2.51	2.5
MON#	0.19	0.18	0.22
EOS#	0.21	0.12	0.19
BAS#	0.1	0.05	0.02
RBC	4.42	4.43	4.42
HGB	13.3	13.1	13.2
MCHC	37.1	36.4	36.7
MCH	30	29.5	29.8
MCV	80.8	80.9	81.1
RDW-CV	17.3	17.3	17.1
RDW-SD	53.2	53.2	53.2
HCT	35.7	35.8	35.8
PLT	194	202	203
MPV	8.7	9.2	9
PDW	16.7	17.9	16.2
PCT	0.169	0.186	0.182
P-LCR	23.1	27.1	24.8

Print Time: 23/07/2024 17:03



# ISHA DIAGNOSTIC LABORATORYL-J QC Data List

File No.: 9

Lot No.: 050524

Expiry date: 28-07-2024

Mode: Closed-WB

Level: L

Date	13-06-2024	13-06-2024	13-06-2024	13-06-2024	13-06-2024	15-06-2024	18-06-2024	19-06-2024	21-06-2024	22-06-2024	23-06-2024	26-06-2024
Time	14:36	14:37	14:39	14:41	14:43	10:50	12:04	09:12	09:48	09:41	11:03	11:37
WBC	2.93	2.93	2.81	3	2.85	2.93	2.88	2.89	2.82	2.88	2.9	2.94
NEU%	45.5	47.8	48.8	47.5	47.4	53.8	43.6	43.6	43.6	44.8	39.5	44.4
LYM%	49	46.3	45.3	47.8	47.8	41	50.2	50.9	50.8	49.9	55.7	50.2
MON%	2.6	2.7	2.6	2.1	2.5	2.3	2.7	2.7	2.4	2.6	2.7	2.2
EOS%	2.7	2.7	2.9	2	2	2.4	3.2	2.2	2.6	2.5	1.8	2.8
BAS%	0.2	0.5	0.4	0.6	0.3	0.5	0.3	0.6	0.6	0.2	0.3	0.4
NEU#	1.33	1.41	1.38	1.42	1.35	1.58	1.26	1.26	1.23	1.29	1.15	1.31
LYM#	1.43	1.35	1.27	1.44	1.36	1.2	1.44	1.47	1.43	1.44	1.61	1.48
MON#	0.08	0.08	0.07	0.06	0.07	0.07	0.08	0.08	0.07	0.07	0.08	0.06
EOS#	0.08	0.08	0.08	0.06	0.06	0.07	0.09	0.06	0.07	0.07	0.05	0.08
BAS#	0.01	0.01	0.01	0.02	0.01	0.01	0.01	0.02	0.02	0.01	0.01	0.01
RBC	2.45	2.46	2.45	2.46	2.45	2.44	2.43	2.43	2.5	2.42	2.48	2.47
HGB	7.3	7	7	7.2	7.1	7.1	7	7.2	7.1	7.1	7.1	7.3
MCHC	38.2	36.5	36.7	37.8	37.4	37.3	36.8	37.9	36.4	37.2	36.6	37.6
MCH	29.7	28.4	28.6	29.5	28.9	29.1	28.7	29.8	28.7	29.2	28.8	29.7
MCV	77.9	77.8	77.9	78	77.2	77.9	78	78.6	78.6	78.7	78.6	79
RDW-CV	18.3	17.8	18.3	16.9	17.2	16.9	17.5	18.1	17.5	17.5	16.9	17.5
RDW-SD	52.1	52.1	53.2	51.1	50	50	51.1	53.2	51.1	52.1	51.1	52.1
HCT	19.1	19.1	19.1	19.1	18.9	19	18.9	19.1	19.6	19.1	19.5	19.5
PLT	62	75	64	67	66	66	58	63	57	71	75	69
MPV	8.6	8.6	8.6	8.8	8.5	8.6	9	8.3	8.7	8.7	9	8.9
PDW	14.3	13.5	17.7	16.2	16.1	15.3	15.4	16.7	15.3	15.3	16.8	15
PCT	0.053	0.064	0.055	0.059	0.057	0.057	0.052	0.053	0.049	0.061	0.067	0.062
P-LCR	21.8	21.3	22.6	23.2	21.4	20.6	24.5	18.2	20.1	22.5	24.5	24.5

Print Time: 10/07/2024 10:02



# ISHA DIAGNOSTIC LABORATORYL-J QC Data List

File No.: 9

Lot No.: 050524

Expiry date: 28-07-2024

Mode: Closed-WB

Level: L

Date	28-06-2024	29-06-2024																		
Time	10:15	11:48																		
WBC	2.95	2.96																		
NEU%	42.9	43.1																		
LYM%	53.5	50.7																		
MON%	2	3.4																		
EOS%	1.3	2.4																		
BAS%	0.3	0.4																		
NEU#	1.26	1.27																		
LYM#	1.58	1.51																		
MON#	0.06	0.1																		
EOS#	0.04	0.07																		
BAS#	0.01	0.01																		
RBC	2.51	2.46																		
HGB	7.1	7.1																		
MCHC	35.9	36.7																		
MCH	28.3	28.8																		
MCV	79	78.4																		
RDW-CV	17.5	17.5																		
RDW-SD	52.1	51.1																		
HCT	19.8	19.3																		
PLT	74	74																		
MPV	8.6	8.9																		
PDW	15	16.9																		
PCT	0.063	0.066																		
P-LCR	21.5	25																		

Print Time: 10/07/2024 10:02



# ISHA DIAGNOSTIC LABORATORYL-J QC Data List

File No.: 10

Lot No.: 050524

Expiry date: 28-07-2024

Mode: Closed-WB

Level: M

Date	13-06-2024	15-06-2024	16-06-2024	19-06-2024	20-06-2024	21-06-2024	23-06-2024	24-06-2024	26-06-2024	27-06-2024	29-06-2024	30-06-2024
Time	14.45	10.52	12.06	09.15	11.43	09.49	11.01	09.48	11.39	09.36	11.49	11.36
WBC	8.54	8.35	8.38	8.78	8.52	8.61	8.61	8.67	8.76	8.54	8.93	8.6
NEU%	55.1	58.7	51.3	49.7	50.5	50.2	53.2	51.5	52.2	48.9	48.6	50.5
LYM%	39.3	36.7	43.7	45.8	45	44.8	41.5	43	42.9	45.6	47.2	43.6
MON%	2.5	2.1	2.3	2	1.9	1.7	2	1.7	2.1	2	1.6	2.4
EOS%	3	2.5	2.5	2.2	2.5	3	2.9	3.3	2.4	3.2	2.3	3.3
BAS%	0.1	0	0.2	0.3	0.1	0.3	0.4	0.5	0.4	0.3	0.3	0.2
NEU#	4.7	4.89	4.3	4.36	4.31	4.31	4.59	4.46	4.57	4.18	4.34	4.34
LYM#	3.36	3.07	3.66	4.02	3.83	3.86	3.57	3.73	3.76	3.89	4.21	3.75
MON#	0.21	0.18	0.19	0.18	0.16	0.15	0.17	0.15	0.18	0.17	0.14	0.21
EOS#	0.26	0.21	0.21	0.19	0.21	0.26	0.25	0.29	0.21	0.27	0.21	0.28
BAS#	0.01	0	0.02	0.03	0.01	0.03	0.03	0.04	0.04	0.03	0.03	0.02
RBC	4.33	4.35	4.43	4.37	4.38	4.37	4.35	4.32	4.41	4.48	4.39	4.43
HGB	13.2	13.2	13.2	13.2	13.3	13.3	13.4	13.3	13.3	13.4	13.3	13.4
MCHC	37.7	37.6	37	37	37.5	37.4	37.8	37.9	37.1	36.8	37.1	37.2
MCH	30.5	30.4	29.9	30.1	30.3	30.3	30.7	30.8	30.2	29.9	30.2	30.2
MCV	81	81	80.8	81.4	80.9	81.2	81.3	81.3	81.6	81.4	81.4	81.2
RDW-CV	16.8	17.3	17.6	17.1	16.8	16.8	17.3	17.3	17.3	17.3	17.3	16.8
RDW-SD	51.1	52.1	53.2	53.2	52.1	52.1	53.2	54.3	53.2	54.3	54.3	53.2
HCT	35	35.3	35.8	35.6	35.4	35.5	35.3	35.2	36	36.5	35.8	36
PLT	191	189	198	201	198	195	196	204	205	212	220	217
MPV	8.8	8.8	9.2	8.7	8.8	9.1	9.1	9	9.1	8.9	8.9	8.9
PDW	17.1	16.7	17.7	16.9	16.8	16.3	17.2	17.8	17.1	16.7	17.9	16.1
PCT	0.168	0.166	0.181	0.175	0.175	0.178	0.179	0.183	0.185	0.189	0.196	0.194
P-LCR	22.8	23.5	27.1	22.4	24.9	26.5	26.5	25.6	25.5	26.2	25.3	25.5

Print Time: 10/07/2024 10:03



# ISHA DIAGNOSTIC LABORATORYL-J QC Data List

File No.: 11

Lot No.: 050524

Expiry date: 28-07-2024

Mode: Closed-WB

Level: H

Date	13-06-2024	16-06-2024	18-06-2024	20-06-2024	22-06-2024	24-06-2024	27-06-2024	28-06-2024	30-06-2024			
Time	14.47	12.08	12.05	11.44	09.43	09.52	09.40	10.13	11.38			
WBC	15.39	15.78	15.49	15.7	15.99	15.4	15.94	15.79	15.79			
NEU%	67	66.6	68.2	65.9	68.7	67.5	68	66.2	64.6			
LYM%	27.2	20.5	26.4	22.2	25.8	27	27.1	28.5	30.3			
MON%	2.8	5.8	3	5.9	2.4	2.3	2.1	2.4	2.3			
EOS%	2.7	3.4	2.2	3.2	3	3.1	2.7	2.7	2.7			
BAS%	0.3	3.7	0.2	2.8	0.1	0.1	0.1	0.2	0.1			
NEU#	10.3	10.5	10.57	10.34	10.98	10.39	10.84	10.45	10.2			
LYM#	4.19	3.24	4.09	3.49	4.13	4.16	4.32	4.5	4.78			
MON#	0.43	0.92	0.46	0.93	0.38	0.35	0.33	0.38	0.36			
EOS#	0.42	0.54	0.34	0.5	0.48	0.48	0.43	0.43	0.43			
BAS#	0.05	0.58	0.03	0.44	0.02	0.02	0.02	0.03	0.02			
RBC	5.52	5.52	5.52	5.54	5.52	5.49	5.49	5.44	5.53			
HGB	17	17	16.9	17.1	17.1	17.1	17	17.1	17.1			
MCHC	37	37.1	36.9	36.9	37.1	37.4	37.1	37.7	36.9			
MCH	30.8	30.8	30.7	30.8	31	31.1	31	31.4	30.9			
MCV	83.4	83.1	83.2	83.5	83.5	83.2	83.6	83.3	83.5			
RDW-CV	16.9	16.9	17.4	16.7	17.2	16.9	17.4	17.4	17.4			
RDW-SD	54.3	53.2	55.3	53.2	54.3	54.3	55.3	55.3	54.3			
HCT	46	45.9	45.9	46.2	46.1	45.7	45.9	45.4	46.2			
PLT	333	380	371	4	382	392	406	402	390			
MPV	8.9	8.9	8.8	8.9	9	9	8.9	8.8	8.9			
PDW	17.6	17	16.5	17.5	17.1	17	16.6	16.4	16.8			
PCT	0.296	0.34	0.327	0.004	0.343	0.354	0.363	0.353	0.348			
P-LCR	24.4	25.2	23.1	25.1	24.7	25.6	25.4	23.2	24.4			

Print Time: 10/07/2024 10:04

