

BHARATH INSTRUMENTS

OFFICE : #36/239, Angappa Naicken Street, Mannady, Chennai - 600 001.



CALIBRATION CERTIFICATE

BI:CC.No:0386/2024

Name & Address

: ICTC,
PHC HUTBAY.

Master Instrument Reference

: Tachometer, LT Make, DT-2236 Model,
Sl.No.:132076, Range: 5 to 99999 RPM, Calibrated at
PCCPL, Chennai - 600 008. Report No.:PC-12-22/ATL/5194-01,
Dtd: 05-12-2022. Valid upto 04-12-2024. (NABL)

Test Instrument Reference

: CENTRIFUGE, REMI Make, Model: C-854/8,
Range: 0 to 3500RPM, Sl. No.: ZIAN-02844 (BI:0386/2024),
Identification No.: PHC/CENT/ICTC/CENTR-01.

Ambient Temp. When tested : 30°C

TEST READINGS

Master Instrument Readings	Test Instrument Readings	Error (if any)	
		Allowed	Observed
rpm	rpm	± 1 %	rpm
502	500	"	-2
1005	1000	"	-5
2018	2000	"	-18
3023	3000	"	-23

Remarks: The above instruments well within the specified accuracy

Calibrated by : *c.vath*
Calibrated on : 30-01-2024
Next Calibn. Due on : 29-01-2025



For BHARATH INSTRUMENTS
A.k
Signatory