


Analyte	Instrument	Result Value	Standard Unit
X Blood (urine)	Manual	Trace	-
X Nitrite (urine)	Manual	Absent	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⊙ Delayed Result Entry			

Problem Classification: Transcription Error done by Technician.

Corrective Action: Technician sign and supervisor sign implemented on raw data record

Reviewed by:  15/07/2024

Dated: 15/07/2024

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	5	5,5,5,6
@ Specific Gravity	-	1.030	1.020,1.025,1.030
@ Glucose (urine)	-	2+	+2,+3,+4
@ Protein (Urine)	-	1+	+1,+2,+3
@ Ketone (urine)	-	Negative	+1,Neg,Tr
X Blood (urine)	-	Trace	+1,+2,+3
@ Bilirubin (urine)	-	Negative	+1,Neg
@ Urobilinogen (urine)	-	Normal	+1,Nor
X Nitrite (urine)	-	Absent	Pre

Legend @ : Acceptable  
 @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⊙ : Delayed Result Entry

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	2
EQAS Score Clinical Path - Urine Routine Chemistry	77.78 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kirol Road, Kurla (W),  
 Mumbai - 400070



Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Crystal	-	Triple phosphate crystal	Triple phosphate crystal
Legend @ : Acceptable @ : Acceptable X : Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry			
		Total Parameters	1
		Not Evaluated Parameters	0
		Evaluated Parameters	1
		Outlier Parameters (X)	0
		EQAS Score Clinical Path - Urine Sediment Analysis Image	100.00 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kirod Road, Kurla (W),  
 Mumbai - 400070