

Analyte	Instrument	Result Value	Standard Unit
X Widal test - Typhi O antigen (TO)	Manual	1:80	-
X Widal test - Typhi H antigen (TH)	Manual	1:80	-

Legend @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ Delayed Result Entry

Problem Classification: Transcription Error.

Corrective Action: Technician sign and supervisor sign implemented on raw data record.

Reviewed by: [Signature]
 15/07/24

Dated: 15/07/24

Instrument : J Mitra

Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ HIV Ab	S/CO	Non-reactive	Non-reactive

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
X Widal test - Typhi O antigen (TO)	-	1:80	No Agglutination
X Widal test - Typhi H antigen (TH)	-	1:80	No Agglutination
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	No Agglutination

Legend @ : Acceptable
@ : Acceptable
X : Unacceptable
: Not Evaluated
⊙ : Delayed Result Entry

Total Parameters	8
Not Evaluated Parameters	0
Evaluated Parameters	8
Outlier Parameters (X)	2
EQAS Score Immunoassay Extended	75.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirod Road, Kurla (W),
Mumbai - 400070

Analyte	Instrument	Result Value	Standard Unit
X Widal test - Typhi O antigen (TO)	Manual	1:80	-
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⊙ : Delayed Result Entry			

Problem Classification: Transcription Error

Corrective Action: Technician Sign and Supervisor Signature implemented as per date record,

Reviewed by: [Signature]
 15/07/2024

Dated: 15/07/2024

Instrument : J Mitra

Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Reactive	Reactive
@ HCV	IU/mL	Non-reactive	NR
@ HIV Ab	S/CO	HIV-1/HIV-2 Ab Reactive	HIV-1/HIV-2 Ab Reactive

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
X Widal test - Typhi O antigen (TO)	-	1:80	No Agglutination
@ Widal test - Typhi H antigen (TH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	No Agglutination
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	No Agglutination

- Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry

Total Parameters	8
Not Evaluated Parameters	0
Evaluated Parameters	8
Outlier Parameters (X)	1
EQAS Score Immunoassay Extended	87.50 %

Dr Puneet Kumar Nigam


Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070

Analyte	Instrument	Result Value	Standard Unit
# HIV Ab	J Mitra	Non-reactive	-

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
⊙ Delayed Result Entry

Problem Classification: No outlier

Corrective Action: No outlier

Reviewed by:  11/07/2024

Dated: 11/07/2024

Instrument : J Mitra

Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Reactive	Reactive
@ HCV	IU/mL	Reactive	Reactive
# HIV Ab	S/CO	Non-reactive	

Legend @ : Acceptable	Total Parameters	3
@ : Acceptable	Not Evaluated Parameters	1
X : Unacceptable	Evaluated Parameters	2
# : Not Evaluated	Outlier Parameters (X)	0
⊙ : Delayed Result Entry	EQAS Score Immunoassay Extended	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070