

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RM
@ Erythrocyte (RBC) Count	mill/cu.mm	2.46	2.47	-0.09	0.
@ Total Leucocytes (WBC) count	/c.mm	2520.00	2649.14	-0.52	-0.
@ Haemoglobin	g/dL	6.80	6.75	-0.66	-0.
@ MCV (Mean Corpuscular Volume)	fL	104.90	106.32	-0.39	-0.
@ MCH (Mean Corpuscular Hb)	pg	27.50	27.25	0.16	-0.
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	26.30	25.58	0.44	-0.
@ Platelet Count	1000/ μ L	168.00	158.56	1.25	0.
@ Hematocrit	%	25.80	26.17	-0.67	-0.
@ RDW CV%	%	14.30	18.60	-1.44	-1.

Legend @ : Acceptable

! : $2.0 < z \text{ score} < 3.0$ - Warning

X : $z \text{ score} \geq 3$ - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	!
Not Evaluated Parameters	!
Evaluated Parameters	!
Outlier Parameters (X)	!
EQAS Score Haematology - CBC	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroi Road, Kurla (W).
 Mumbai - 400070

Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD HOSHIARPUR (1050)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
---------	------------	--------------	---------------	---------

No Outlier

Legend @ : z score \leq 2.0 - Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score \geq 3 - Unacceptable
: Not Evaluated
⊙ : Delayed Result Entry
* : Not considered for evaluation.

Problem Classification: NA.

Corrective Action: No - outlier.

Reviewed by: [Signature]

Dated: 11/07/24

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RM
@ Erythrocyte (RBC) Count	mill/cu.mm	2.79	2.73	0.10	0.
@ Total Leucocytes (WBC) count	/c.mm	5300.00	7767.57	-1.89	-0.
@ Haemoglobin	g/dL	8.20	8.06	-0.36	-0.
@ MCV (Mean Corpuscular Volume)	fL	115.40	117.70	-0.31	-0.
@ MCH (Mean Corpuscular Hb)	pg	29.30	29.56	-0.09	-0.
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	25.40	24.78	0.46	-0.
@ Platelet Count	1000/ μ L	287.00	271.83	0.99	0.
@ Hematocrit	%	32.30	31.93	-0.45	-0.
! RDW CV%	%	13.50	18.09	-2.35	-1.

Legend @ : Acceptable

! : $2.0 < z \text{ score} < 3.0$ - Warning

X : $z \text{ score} \geq 3$ - Unacceptable

: Not Evaluated

⊙ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	!
Not Evaluated Parameters	!
Evaluated Parameters	!
Outlier Parameters (X)	!
EQAS Score Haematology - CBC	100.00 %

Dr Puneet Kumar Nigam

Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirod Road, Kurla (W),
 Mumbai - 400070

Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD HOSHIARPUR (1050)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 02

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Sco
I RDW CV%	Horiba Yumizen H550	13.50	%	-2.0

- Legend
- @ : z score \leq 2.0 - Acceptable
 - I : 2.0 < z score < 3.0 - Warning
 - X : z score \geq 3 - Unacceptable
 - # : Not Evaluated
 - ⊙ : Delayed Result Entry
 - * : Not considered for evaluation.

Problem Classification: NA

Corrective Action: Nat - outlier.

Reviewed by: [Signature]

Dated: 11/7/24