

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
X Widal test - Typhi O antigen (TO)	-	1:160	No Agglutination
X Widal test - Typhi H antigen (TH)	-	1:160	No Agglutination
X Widal test - ParaTyphi A-H antigen (AH)	-	1:80	No Agglutination
@ RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ Widal test - ParaTyphi B-H antigen (BH)	-	No Agglutination	No Agglutination
@ Dengue IgM	RU/ml	Negative	Negative
@ Dengue IgG	RU/ml	Negative	Negative

Instrument : J Mitra

Analyte	Standard Unit	Result Value	Accepted Value
@ HBsAg	IU/mL	Non-reactive	Non-reactive
@ HCV	IU/mL	Non-reactive	Non-reactive
@ HIV Ab	S/CO	Non-reactive	Non-reactive

Legend @ : Acceptable  
@ : Acceptable  
X : Unacceptable  
# : Not Evaluated  
⊙ : Delayed Result Entry

Total Parameters	1
Not Evaluated Parameters	0
Evaluated Parameters	1
Outlier Parameters (X)	0
EQAS Score Immunoassay Extended	72.73 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kirod Road, Kurla (W),  
Mumbai - 400070

Analyte	Instrument	Result Value	Standard Un
X Widal test - Typhi O antigen (TO)	Manual	1:160	-
X Widal test - Typhi H antigen (TH)	Manual	1:160	-
X Widal test - ParaTyphi A-H antigen (AH)	Manual	1:80	-

Legend @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⊙ Delayed Result Entry

Problem Classification: Integruato Issue.

Corrective Action: ILC done. and ILC found satisfactory.

Reviewed by: Pooja

Dated: 10/07/24

Instrument : J Mitra

Analyte	Standard Unit	Result Value	Accepted Value
X HBsAg	IU/mL	Non-reactive	Reactive
X HIV Ab	S/CO	Non-reactive	HIV-1/HIV-2 Ab Reactive
@ HCV	IU/mL	Non-reactive	Non-reactive

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
X Widal test - Typhi O antigen (TO)	-	1:80	No Agglutination
X Widal test - Typhi H antigen (TH)	-	1:80	No Agglutination
X Widal test - ParaTyphi B-H antigen (BH)	-	1:160	No Agglutination
● RPR (Syphilis Serology)	-	Non-reactive	Non-reactive
@ Dengue NS 1	RU/ml	Negative	Negative
@ Widal test - ParaTyphi A-H antigen (AH)	-	No Agglutination	No Agglutination
@ Dengue IgM	RU/ml	Negative	Negative
@ Dengue IgG	RU/ml	Negative	Negative

Legend @ : Acceptable  
 ● : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⊙ : Delayed Result Entry

Total Parameters	1
Not Evaluated Parameters	1
Evaluated Parameters	1
Outlier Parameters (X)	1
EQAS Score Immunoassay Extended	54.55 %

*Dr Puneet Kumar Nigam*

Dr Puneet Kumar Nigam  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070

Analyte	Instrument	Result Value	Standard L
X HBsAg	J Mitra	Non-reactive	-
X HIV Ab	J Mitra	Non-reactive	-
X Widal test - Typhi O antigen (TO)	Manual	1:80	-
X Widal test - Typhi H antigen (TH)	Manual	1:80	-
X Widal test - ParaTyphi B-H antigen (BH)	Manual	1:160	-

Legend @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⊙ : Delayed Result Entry

Problem Classification: Sample integration issue.

Corrective Action: JLC done and acceptable.

Reviewed by: [Signature]

Dated: 18/07/24

**Metropolis Healthcare Ltd., EQAS**

KRSNAA DIAGNOSTICS LTD HOSHIARPUR (1050)

Outlier And Analyte Summary Report

Result Details For Cycle No 240104 and Sample No 03

Report Date : 10/07/2024

Instrument : J Mitra

Analyte	Standard Unit	Result Value	Accepted Value
# HIV Ab	S/CO	Non-reactive	
@ HBsAg	IU/mL	Reactive	Reactive
@ HCV	IU/mL	Reactive	Reactive

Legend @ : Acceptable

@ : Acceptable

X : Unacceptable

# : Not Evaluated

⊙ : Delayed Result Entry

Total Parameters	:
Not Evaluated Parameters	:
Evaluated Parameters	:
Outlier Parameters (X)	:
EQAS Score Immunoassay Extended	100.00 %

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroli Road, Kurla (W),  
 Mumbai - 400070

# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD HOSHIARPUR (1050)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 03

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Un
HIV Ab	J Mitra	Non-reactive	-

- Legend
- @ : Acceptable
  - X : Unacceptable
  - # : Not Evaluated
  - ⊙ : Delayed Result Entry

Problem Classification: NA.

Corrective Action: Not evaluated

Reviewed by: Pooja

Dated: 11/07/24