



HORIBA India Private Ltd.

246, Okhla Industrial
Estate, Phase - III 110
020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2024-2025/1469

20th July 2024.

CALIBRATION CERTIFICATE

This is to certify that the Haematology Analyzer **ABX Yumizen H550** bearing serial number: **204YAXH03682** installed **Krasna Diagnostic centre Hoshiarpur** was Calibrated on **20th July 2024.**

Calibrator: ABX MINOCAL

Lot No.: CX497

Expiry Date: 05th August 2024.

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 19th July 2025.



Shrish Dixit

(Head- Products & Marketing)

For **Horiba India Pvt. Ltd.**