



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2024

LAB NO: 19896

LAB NAME: **GAUTAM LABDHI MEDICAL & RESEARCH CHARITAE**

CONSTITUENT GROUP

Chemistry I

Date of Result Entered: 10/06/2024

Date of Result Published: 27/06/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	D.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	279	1.09	23.28	0.25	1.17 mg/dL	0.32	0.03

u* - method of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

LAB ADDRESS

GAUTAM LABDHI MEDICAL & RESEARCH
CHARITABLE TRUST,
SHOP NO.106, 107, ASHOKA PLAZA, NEXT TO
HOTEL HYAT, PUNE VIMANNAGAR ROAD,

Coordinator Contact Details:

E-Mail: clinqc@cmcvellore.ac.in
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential.
CMC EQAS does not sub contract any components.
***** End of Report *****

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.