

# Ashuma Diagnostic Solutions (OPC) Pvt. Ltd.

## CALIBRATION RECORD FOR THE CENTRIFUGE MACHINE

Institute Name:	Deep Clinical Lab			No. 13/2024-25	
Institute Address:	Karvenafgar Pune 52				
Instrument:	Centrifuge	Make:	Euro	Instrument Model:	
Instrument Serial No:	Brushless Centrifuge				
Instrument Calibration Date:	03.04.2024				
Instrument Calibration Due Date:	02.04.2025				

### A) Speed Measurement

Instrument Used:	Digital Techometer	Serial No:	LH2590028	Make:	Metrix
------------------	--------------------	------------	-----------	-------	--------

Serial No	RMP Set by Knob	RMP on Display	RPM on Techometer	Tolerance RPM
1	1000	1000	1010.2	-10.20
2	2000	2000	2005	-5.00
3	2500	2500	2508	-8.00
4	3000	3000	3007	-7.00
5	3500	3500	3495.0	5.00
6	4000	4000	3990	10.00

Remark: Satisfactory

Performed By: Ashutosh

Checked By: Ashutosh Mandavgane

Date: 03.04.2024

Date: 03.04.2024

### B) Time Measurement

Instrument Used:	Stop Watch	Serial No:		Make:	
------------------	------------	------------	--	-------	--

Serial No	Time Set (Min)	Time on Display	Time on Stop Watch (Min)	Tolerance (Min)
1	5	5	5	0.00
2	10	10	10	0.00
3	15	15	15	0.00

Remark: Satisfactory

Performed By: Ashutosh

Checked By: Ashutosh Mandavgane

Date: 03.04.2024

Date: 03.04.2024

