



# Metropolis Healthcare Ltd., EQAS

TRANSCARE LABORATORIES PVT. LTD. (686)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Creatinine	Mindray BS-240	1.30	mg/dl	2.09
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_


**Dated:** \_\_\_\_\_

Instrument : Mindray BS-240

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	34.90	35.63	-0.27	0.45
! Creatinine	mg/dl	1.30	0.91	2.09	0.32
@ Bilirubin Total	mg/dl	0.20	0.23	-0.13	-0.02
@ Bilirubin Direct	mg/dl	0.10	0.14	-1.30	0.46
@ Total Protein	g/dL	7.70	7.25	1.20	0.17
@ Albumin	g/dL	3.80	4.14	-1.99	5.93
@ Uric Acid	mg/dl	4.43	4.50	-0.10	-0.92
@ Calcium	mg/dl	8.74	8.84	-0.18	-0.66
@ Cholesterol	mg/dl	169.40	172.58	-0.24	-0.48
@ Triglycerides	mg/dl	142.90	145.14	0.13	0.06
@ ALT	U/L	17.00	19.99	-0.48	-0.42
@ AST	U/L	31.00	34.06	-0.24	-0.92
@ ALP	U/L	86.00	85.11	0.33	0.59

Legend @ : Acceptable  
! : 2.0 < z score < 3.0 - Warning  
X : z score ≥ 3 - Unacceptable  
# : Not Evaluated  
⌚ : Delayed Result Entry  
\* : Not considered for evaluation.

Total Parameters	13
Not Evaluated Parameters	0
Evaluated Parameters	13
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kiroi Road, Kurla (W),  
Mumbai - 400070