

QC Result

Control	Lot No.	Chemistry	Lot No.	Unit	Measured	Set		Flag	Run Date/Time
					Result	Mean	SD		
RANDOX 2	1593UN	Ca		mg/dL	8.17	8.7	0.44		10-04-2024 13:41:07
RANDOX 2	1593UN	Glu-G		mg/dL	102.9	114	8.35		10-04-2024 13:48:52
RANDOX 2	1593UN	T-bil-D II		mg/dL	1.75	1.73	0.19		10-04-2024 13:49:23
RANDOX 2	1593UN	D-bil-D II		mg/dL	1.12	1.21	0.13		10-04-2024 13:49:22
RANDOX 2	1593UN	ALT		U/L	30.6	32	3		10-04-2024 13:49:07
RANDOX 2	1593UN	AST		U/L	28.1	32	3.5		10-04-2024 13:49:22
RANDOX 2	1593UN	ALP		U/L	180.1	193	14.5		10-04-2024 13:49:37
RANDOX 2	1593UN	γ-GT		U/L	43.5	46	3.5		10-04-2024 13:49:52
RANDOX 2	1593UN	TP		g/dL	6.6	5.96	0.6		10-04-2024 13:50:53
RANDOX 2	1593UN	ALB		g/dL	3.7	4.21	0.32		10-04-2024 13:47:07
RANDOX 2	1593UN	TG		mg/dL	95.7	102	8.35		10-04-2024 13:50:37
RANDOX 2	1593UN	TC		mg/dL	153.5	162	10.5		10-04-2024 13:47:37
RANDOX 2	1593UN	CREA-S		mg/dL	1.33	1.42	0.14		10-04-2024 13:52:37
RANDOX 2	1593UN	UA		mg/dL	5.26	5.85	0.38		10-04-2024 13:52:52
RANDOX 2	1593UN	UREA		mg/dL	44.19	43.8	3.3		10-04-2024 13:51:37
RANDOX 2	1593UN	Fe		μg/dL	105.7	108	9.55		10-04-2024 13:53:53