

## CERTIFICATE OF CALIBRATION

<b>Certificate No</b>	CAL-AR-2024-04-0300		
<b>Date of Receipt</b>	17-04-2024	<b>Issue Date</b>	23-04-2024

<b>Issued By</b>	Trimed Solutions (I) Private Limited		
<b>Address</b>	2nd Floor, Jyothi Elegance, Kavuri Hills, Phase-III, Madhapur, Hyderabad-500081		

<b>Health Facility</b>	Khando Dorwa Sangmo DH, Tawang		
<b>State</b>	Arunachal Pradesh	<b>District</b>	Tawang

### Medical Equipment Details

<b>ME Number</b>	ARZ1TAWDH0090011	<b>Manufacturer</b>	Electra
<b>ME Name</b>	Centrifuges, Tabletop (Centrifuge)	<b>Model</b>	NA
<b>Department</b>	Laboratory-Na	<b>Serial No</b>	NA
<b>Range</b>	300-3000RPM	<b>Resolution</b>	Not Available
<b>Calibrated at</b>	On site	<b>Condition of Equipment</b>	Visually good
<b>Date of Calibration</b>	17-04-2024	<b>Calibration Due</b>	16-04-2025

### Environmental Conditions

<b>Temperature in °C</b>	20°C	<b>Relative Humidity in % Rh</b>	58%
--------------------------	------	----------------------------------	-----

### Details of Master Equipment used for Calibration

No	Description	Equipment ID	Certificate Number	Date of Calibration	Calibration Due
1	Digital Tachometer	TEI-TE-0015	TSPL/23-24/MD-00162	15-11-2023	14-11-2024

### Calibration Standards



<b>Applicable Specification</b>	By Using Digital Tachometer By Direct Method	<b>Calibration Procedure</b>	TSPL-CAL-11
---------------------------------	---	------------------------------	-------------

### Performance Result

No	Parameter	Unit	Set value	Measured Value	Acceptable range	Remarks
1	Speed	RPM	1000	1050	900 To 1100	Pass
2		RPM	2000	2110	1800 To 2200	Pass
3		RPM	3000	3150	2700 To 3300	Pass

<b>Conditions of Issue of Report</b>	
1.	The standards used are traceable to National and International Standards through unbroken chain of Accredited Laboratories.
2.	Results reported are valid at the time of & under the stated condition of the measurements.
3.	Any error in the certificate should be brought to our knowledge within 30 days from the date of the certificate.
4.	This calibration report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.
5.	Decision Rule - Simple Acceptance where "PASS" indicates measured value is within/Equal to the Tolerance mentioned by Manufacturer specification and "FAIL" indicates measured value is out of Tolerance without considering the measurement Uncertainty.

<b>Calibration Status</b>	<b>Pass</b>	Measured Values are within the Specified Limit
---------------------------	-------------	--

<b>Calibrated By</b>  <b>Midhun CM</b>	<b>Authorised Signatory</b>  <b>Rahul Kumar Kashyap</b>
---	---