



HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

HIN/MED/2024-2025 /105567201
19THAPRIL 2024

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer Horiba YUMIZEN H1500 bearing serial number 302M1XH01076 installed at KRSNAA DIAGNOSTICS LTD UNA as calibrated on 19TH APRIL 2024.

Calibrator : ABX MINOCAL
Lot No. : CX493
Expiry Date : 05th May 2024

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 18TH OCT 2024



Shrish Dixit

(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**

246, Okhla Industrial Estate Phase-
III, New Delhi 110020, India
Tel : +91 (11) 4546 5000 / 46695001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC1M232

