



UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY



CALIBRATION CERTIFICATE

FormatNo.:UCCS-QF-7.8.1	CertificateNo.:UCCSCC20221207-015	ULRNo.:CC328722000005405F
Group:Acceleration&Speed		Discipline:Mechanical

Customer'sName&Address M/s. PCR HEALTHCARE PVT.LTD RZ11 MAIN ROAD INDRA PARK PALAM COLONY NEW DELHI	PageNo.: 1of1	CalibrationPerformed:AtSite
	DateofReceipt:11.11.2023	
	ServiceRequestNo.:UCCSSRN20221207-004	
	CertificateIssueDate:11.11.2023	
	DateofCalibration:12.11.2024	
NextDueDate(SuggestedbyCustomer):10.11.2024		

EquipmentDescription(UUC)&Identification					
NameoftheInstrument:DigitalRPMIndicator (Centrifuge Machine)			Make: Remi		
Range: Upto5000			Model:r-R-23		
LeastCount:10RPM			Sr.No.:ZHEN-15546		
ConditionofInstrument: Ok			I.D.No.:-----		
MasterEquipmentUsed			Location:LAB		
Sr.No.	InstrumentName	Range&I.D.No.	CalibrationAgency	CalibrationCertificateNo.	ValidUpto
1	DigitalTachometer	2.5To99999RPM&U CCS/CAL/014	AsianTechnology	AT22000000095	06-03-2024
EnvironmentalConditions			CalibrationProcedure	CalibrationReferenceStandard	
Temperature: 25±5°C, Humidity:50±10%RH			UCCS03SOP033	----	

CALIBRATION RESULTS

Sr.No.	Set UUC Value(R PM)	Observed Ref. Standard Value(RPM)	Expanded Uncertainty(±)
1	500	502	3% of rdg
2	1000	1005	
3	1500	1509	
4	2000	2015	
5	3500	3510	
6	5000	5003	

EndofCertificate

Amit
Calibrated By
Amit Kumar
Calibration Engineer

Authorized
Personnel
Approved by
Vikash Kumar
Technical Manager

Working : W.Z. -1390/Z-2, 1st Floor, Nangal Raya Extn., Pankha Road, New Delhi-110046
Regd. Office : R.Z.-56A, Indra Park, Gali No.25, Palam Colony, New Delhi-110045
M.: 9958445875, 9313768426, 9999761885 E-Mail: uniquecalibrationservices@gmail.com

