




CALIBRATION CERTIFICATE

Instrument Name : Beckman Coulter Bio Chemistry Analyzer AU480.
Instrument Serial Number : 2015084404
Installed at : MORRIS MATHIAS HOSPITAL , NAGERCOIL
Calibration Date : 05/12/2023
Next Calibration Due on : 04/12/2024

Photometric Check

Requirement : All O.Ds be less than 1.5

Wavelength	OD	Result
340	1.0458	Passed
380	0.8206	Passed
410	0.7885	Passed
450	0.8235	Passed
480	0.6409	Passed
520	0.6471	Passed
540	0.7323	Passed
570	0.8452	Passed
600	0.9415	Passed
660	0.7239	Passed
700	0.6857	Passed
750	0.6178	Passed
800	0.5740	Passed


S. Ashwin Kumar
Senior Service Engineer