



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 2382

Distribution No.: 163-F

Month/Year: March/2024

Instrument ID: Transasia

Model Name.: SYSMEX KX-21

Serial No.: B-2817

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra (Prof. & Head), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : info@ishtmaiimseqap.com

Date of issue & status of the report: 10-05-2024[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	6.3	6.3	12.6	12.8	0.053	-0.12	0	0.1	0.008	-0.96
RBC x10 ⁶ /µl	1	4.18	4.12	8.3	8.41	0.007	-0.55	0.06	0.04	0.002	0.54
Hb g/dl	1	13.3	13.2	26.5	26.8	0.020	-0.51	0.1	0.1	0.008	0.00
HCT%	1	40.3	39.7	80	84.5	0.183	-0.82	0.6	0.4	0.024	0.54
MCV-fl	1	96.4	96.4	192.8	200.9	0.360	-0.61	0	0.2	0.019	-0.67
MCH-Pg	1	32.3	31.6	63.9	63.9	0.060	0.00	0.7	0.3	0.018	1.35
MCHC-g/dl	1	33.5	32.8	66.3	63.2	0.126	0.66	0.7	0.3	0.013	1.35
Plt. x10 ³ /µl	1	181	175	356	367	1.134	-0.35	6	4	0.246	0.45
Retic %	2	11	10	21	19	0.335	0.19	1	0.6	0.044	0.77

P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs= , Poly=54 L=16, E=7, Mono/Promono=13 , B1= P.M.=, Mye=4, Meta=6, Other=
RBC Morphology	3	Poly: 64 - 75, Lympho: 6 - 12, Myelo: 4 - 9, Meta: 2 - 6, Eosino: 2- 5, Mono: 1-5, Promyelo/Blast/Baso: 0-5
Diagnosis	3	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Mild: Poikilocytosis, Macrocytes, Tear drop cells
		LEUKOCYTOSIS
		Chronic Myeloid Leukemia (Chronic Phase)

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 163--F	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	345	341	89.15	88.27	4.4	6.74	6.45	4.99
RBC x10⁶/µl	1	345	345	85.51	88.7	7.83	4.93	6.66	6.37
Hb g/dl	1	345	345	85.8	88.41	7.54	4.93	6.66	6.66
HCT%	1	345	341	94.72	91.5	3.23	4.99	2.05	3.51
MCV-fl	1	345	341	98.53	86.51	1.47	5.28	0	8.21
MCH-Pg	1	345	341	87.68	91.79	6.74	3.23	5.58	4.98
MCHC-g/dl	1	345	341	96.19	91.79	3.23	2.64	0.58	5.57
Plt. x10³/µl	1	345	340	90.29	93.24	4.71	3.82	5	2.94
ReticCount%	2	345	288	94.79	82.64	3.82	11.46	1.39	5.90
PS Assessment	3	345	274	Satisfactory :67.84%, Borderline Sat. :24.635, Unsatisfactory :7.53%					

***Comments:**

1). Among Lab (EQA) : PS Diagnosis partially correct, remaining results acceptable

2). Within Lab (IQA) : Precision acceptable.

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ± 2 : Acceptable, Z score ± 2 to ± 3 :Warning Signal, Z score $> \pm 3$: Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ± 2 " are texted in green colour. Z score value between " ± 2 to ± 3 " are texted in orange colour. Z score value $> \pm 3$ are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra (Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----