

**Date: 09-07-2024**

## **CALIBRATION CERTIFICATE**

This is to state that the Haematology Analyzer Model **HORIBA Yumizen H500** bearing serial no: **007YOXH03604** installed at **SCHIEFFELIN INSTITUTE OF HEALTH – RESEARCH & LEPROSY CENTRE, KARIGIRI** was calibrated with Calibrator ABX Minocal on **9<sup>th</sup> July 2024** as per the standard operating procedure.

Subsequently the controls were processed, and the results were found in respective ranges. The instrument was Calibrated successfully.

Calibrator Used: **ABX MINOCAL**  
Lot No.: **CX 496**  
Expiry Date: **05-08-2024**  
Next Calibration Due Date: **08-07-2025.**

**Performed By**



(Logesh. A)  
Sr. Engineer– Customer Support  
**Horiba India Pvt Ltd.,**