

Sensa Core Medical Instrumentation Pvt Ltd

Plot No.3,EPIP Zone,Pashamylaram, Medak District-502307. www.sensacore.com

Instrument Name: ST-200 CL PLUS Electrolyte Analyzer

Date of Testing : 10/07/2024 Valid up to 09/07/2025

Instrument Sr.No.: 10876

Calibration Slope :

S.NO	Na+	K+	iCa	Cl
Mean	66.49	67.24	-	49.91
Ranges	mV/decade	mV/decade	mV/decade	mV/decade
Status	Ok	Ok	-	Ok

Quality Control (I) : Lot:UQC2311B

S.NO	Na+	K+	iCa	Cl
Mean	125.4	2.18	-	79.4
Ranges	mmol/L	mmol/L	mmol/L	mmol/L
Status	Ok	Ok	-	Ok

Quality Control (II) : Lot:UQC2311B

S.NO	Na+	K+	iCa	Cl
Mean	140.9	3.91	-	96.1
Ranges	mmol/L	mmol/L	mmol/L	mmol/L
Status	Ok	Ok	-	ok

Quality Control (III) : Lot:UQC2311B

S.NO	Na+	K+	iCa	Cl
Mean	174.2	6.37	-	131.7
Ranges	mmol/L	mmol/L	mmol/L	mmol/L
Status	Ok	Ok	-	Ok

TESTED BY:

