

Instrument : Roche cobas c 311

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	70.50	74.39	-0.56	-0.14

  

Legend @ : Acceptable	Total Parameters	1
! : 2.0 < z score < 3.0 - Warning	Not Evaluated Parameters	0
X : z score ≥ 3 - Unacceptable	Evaluated Parameters	1
# : Not Evaluated	Outlier Parameters (X)	0
⊙ : Delayed Result Entry	EQAS Score Biochemistry	100.00 %
* : Not considered for evaluation.		

*Dr Puneet Kumar Nigam*

**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kirool Road, Kurla (W),  
 Mumbai - 400070