

# Sensa Core Medical Instrumentation Pvt Ltd

Plot No.3,EPIP Zone,Pashamylaram, Medak Distric-502307. [www.sensacore.com](http://www.sensacore.com)

## Instrument Name: ST-200CL PLUS Na/K/Cl Electrolyte Analyzer

Date of Testing : 26/12/2023

Due Date:25/12/2024

Instrument Sr.No.: **ST-10745**

### Calibration Slope :

S.NO	Na+	K+	Cl
Mean	63.60	69.93	56.01
Ranges	35-85 mV/decade	35-85 mV/decade	35-85 mV/decade
Status	OK	OK	OK

### Quality Control (I) :

S.NO	Na+	K+	Cl
Mean	120.4	2.03	76.7
Ranges	111 – 129 mmol/L	1.50 – 2.50 mmol/L	71-89 mmol/L
Status	OK	OK	OK

### Quality Control (II) :

S.NO	Na+	K+	Cl
Mean	142.9	3.98	100.2
Ranges	135 – 149 mmol/L	3.60 – 4.60 mmol/L	93-107 mmol/L
Status	OK	OK	OK

### Quality Control (III) :

S.NO	Na+	K+	Cl
Mean	162.7	5.96	121.2
Ranges	153 – 171 mmol/L	5.50 – 6.50 mmol/L	111-129 mmol/L
Status	OK	OK	OK

For, YUCCA DIAGNOSTICS.

Authorized signatory with Stamp

Date: 26/12/2023

