



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

**EQAP CODE No. :** 4347

**Distribution No.:** 163-K

**Month/Year:** April/2024

**Instrument ID:** HORIBA MEDICAL

**Model Name.:** YUMIZEN H550

**Serial No.:** 211YADH04052

**Name & Contact No. of PT Co-ordinator:** Dr. Manoranjan Mahapatra ( Prof. & Head), Hematology, AIIMS, Delhi, Tel: 9013085730 , E-Mail : info@ishtmaiimseqap.com

**Date of issue & status of the report:** 19-06-2024[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	5.37	5.31	10.68	8.55	0.033	2.54	0.06	0.1	0.006	-0.45
RBC x10 <sup>6</sup> /µl	1	4.67	4.62	9.29	9.1	0.010	0.78	0.05	0.04	0.003	0.19
Hb g/dl	1	13.6	13.6	27.2	25.5	0.030	2.08	0	0.1	0.007	-1.35
HCT%	1	56.9	56.2	113.1	79.2	0.170	7.79	0.7	0.4	0.023	0.81
MCV-fl	1	121.8	121.7	243.5	174	0.314	8.47	0.1	0.3	0.022	-0.54
MCH-Pg	1	29.5	29.5	59	56.2	0.067	1.73	0	0.2	0.016	-0.90
MCHC-g/dl	1	24.8	23.2	48	64.35	0.161	-3.89	1.6	0.3	0.019	4.38
Plt. x10 <sup>3</sup> /µl	1	211	204	415	179.5	1.782	4.81	7	5	0.297	0.45
Retic %	2	10	9	19	16	0.260	0.42	1	0.5	0.036	0.84

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
<b>DLC%</b>	3	Nrbcs=00 , Poly=10 L=89, E=01, Mono/Promono=00 , B1=00 P.M.=00, Mye=00, Meta=00, Other=Many Smudged cells seen. Prolymphocytes.	Lymp: 77-88, Poly: 8-14.25, mono: 1-3, nRBC/Blast/Myelo/Meta/Eosino: 0-5		
<b>RBC Morphology</b>	3	Normocytic Normochromic	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Mild: Microcytic		
<b>Diagnosis</b>	3	Chronic Lymphoproliferative Disorder CLL.	Chronic Lymphoproliferative Disorder/CLL		

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist. 163--K	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
<b>WBC x10<sup>3</sup>/µl</b>	1	303	303	79.21	89.44	5.61	5.28	15.18	5.28
<b>RBC x10<sup>6</sup>/µl</b>	1	303	303	89.44	90.76	5.28	3.96	5.28	5.28
<b>Hb g/dl</b>	1	303	303	90.1	91.42	5.61	3.96	4.29	4.62
<b>HCT%</b>	1	303	302	93.05	89.4	4.97	5.3	1.98	5.3
<b>MCV-fl</b>	1	303	302	92.05	85.1	5.3	7.28	2.65	7.62
<b>MCH-Pg</b>	1	303	302	87.09	92.38	6.95	4.64	5.96	2.98
<b>MCHC-g/dl</b>	1	303	302	93.05	93.38	3.64	3.31	3.31	3.31
<b>Plt. x10<sup>3</sup>/µl</b>	1	303	302	93.71	91.06	3.64	5.63	2.65	3.31
<b>ReticCount%</b>	2	303	251	95.22	86.85	3.19	7.17	1.59	5.98
<b>PS Assessment</b>	3	303	249	Satisfactory :92.74%, Borderline Sat. :3.30%, Unsatisfactory :3.96%					

**\*Comments:**

1). **Among Lab (EQA) : CBC result for HCT, MCV, MCHC & PLT unacceptable, please check calibration/human error. Remaining results acceptable.**

2). **Within Lab (IQA) : Difference in the CBC measurement values for MCHC unacceptable, may be due to random/human error.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

**Note 10:** Reports are kept confidential.

Report authorized by,





Dr. Manoranjan Mahapatra ( Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----

## GAUTAM LABDHI DIAGNOSTIC CENTER

	NON-CONFORMANCE/ Corrective Action & Preventive Action Report
<b>Details of Non-Conformance: -</b> Haematology EQAS of April 2024 2's core of WtC (-3.89), Hg (7.79), MCV (8.47), Platelet (4.81) count was out.	
<b>Root Cause: -</b> Human error - mixing of EQAS sample was not properly mixed.	
<b>Corrective Action :-</b> review all aspects of the test performance. Review the internal quality control results at the time the EQA sample were measured.	
<b>Preventive Action :-</b> (1) Training given to technician about sample mixing & processing (2) we will run EQA sample of haematology 4 times in two days of span & then submit.	
<b>Follow up :-</b> wait for next cycle result.	
<b>Technical Manager :-</b>	
<b>Sign: -</b>  Mrs. Rajni Gaiwad.	
<b>Date: -</b> 21/06/24	
<b>Effectiveness of Corrective Action Taken by -</b> Mrs. Rajni Gaiwad.	
<b>Technical Manager/ Quality Manager:-</b> Dr. Sushil Agarwal.	
<b>Sign: -</b> 	
<b>Date: -</b>	