

Ashuma Diagnostic Solutions (OPC) Pvt. Ltd.

CALIBRATION RECORD FOR THE CENTRIFUGE MACHINE

Institute Name:	Gautam Labdi Diagnostic Center				
Institute Address:	Nagar Road, Pune				
Instrument:	Centrifuge	Make:	Remi	Instrument Model:	R-8C
Instrument Serial No:	ZIKN-41993				
Instrument Calibration Date:	19.06.2024				
Instrument Calibration Due Date:	18.06.2025				

A) Speed Measurement

Instrument Used:	Digital Techometer	Serial No:	LH2590028	Make:	Metrix
------------------	--------------------	------------	-----------	-------	--------

Serial No	RMP Set by Knob	RMP on Display	RPM on Techometer	Tolerance RPM
1		1000	988.0	12.00
2		2000	2005.3	-5.30
3		3000	2980.5	19.50
4		4000	3989.5	10.50
5		-		

Remark: Satisfactory

Performed By: Sachin Dhokane

Checked By: Ashutosh Mandavgane

Date: 19.06.2024

Date: 19.06.2024

B) Time Measurement

Instrument Used:	Stop Watch	Serial No:		Make:	
------------------	------------	------------	--	-------	--

Serial No	Time Set (Min)	Time on Display	Time on Stop Watch (Min)	Tolerance (Min)
	5 min	5 min	5 min	0.00
	10 min	10 min	10 min	0.00
	15 min	15 min	15 min	0.00

Remark: Satisfactory

Performed By: Sachin Dhokane

Checked By: Ashutosh Mandavgane

Date: 19.06.2024

Date: 19.06.2024

