

CALIBRATION REPORT

Calibration Certificate No:	53/2024
Calibration Date:	01.08.2024
Calibration Due Date:	31.09.2025

Laboratory Name: Gautam labdi Diagnostic Center;

Contact Name: Mrs. Rohini

Address: Viman Nagar, Solapur

Description of Product:

Name: Temperature indicator (Fridge-1)
Sr. No: NA
Make: NA
Model: NA
Type: Digital

Range: 2 to 8°C
Resolution: NA
Working Range: As per Manual
Sr. No: NA
Location: Laboratory

Details of Standard used for calibration:

Name: Digital Temperature mater
Make: Mextech
Model: M 288 CTH
Type: Digital
Condition: Ok

Range: -50 to 70°C
Certificate No: ACIS/2023/02/27/001
Certified by: Accutech Calibration,
Calibration Validity: 26/05/2025

Calibration Results (Temperature)

Fridge Reading displayed in °C	Standard Reading in °C	Error in °C
2.8	3.1	0.3

Deep Fridge Reading displayed	Standard Reading in °C	Error in °C
-0.19	-0.21	0.2

Remark: The value measure above is the mean of 5 Readings

End of the Certificate.

