



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD DASUYA (1044)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 02

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Glucose	mg/dl	75.90	71.18	-0.10	-0.39
@ Urea	mg/dl	25.70	26.57	0.04	-0.43
@ Creatinine	mg/dl	0.90	0.99	-0.60	-0.43
@ Bilirubin Total	mg/dl	0.40	0.35	0.71	0.54
@ Bilirubin Direct	mg/dl	0.10	0.14	0.10	0.64
@ Total Protein	g/dL	6.49	7.58	-0.86	-1.05
@ Albumin	g/dL	3.90	4.19	-0.50	-0.53
@ Uric Acid	mg/dl	4.67	4.32	0.11	-0.29
@ Calcium	mg/dl	9.09	8.55	0.36	0.43
@ Cholesterol	mg/dl	169.00	162.79	-0.01	0.15
@ Triglycerides	mg/dl	145.00	135.91	0.39	0.02
@ HDL	mg/dl	25.00	29.24	-1.19	-0.36
@ LDL	mg/dl	88.45	89.65	-0.53	0.13
@ ALT	U/L	6.70	5.92	0.60	0.15
@ AST	U/L	12.24	13.94	-0.75	1.11
@ ALP	U/L	53.21	45.36	-0.15	-0.68

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	109.57	105.34	0.78	0.39
@ Sodium	mmol/L	123.62	133.36	-0.83	-0.32
@ Potassium	mmol/L	4.98	5.13	-0.39	-0.09

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥3 - Unacceptable

# : Not Evaluated

⊙ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	19
Not Evaluated Parameters	0
Evaluated Parameters	19
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD DASUYA (1044)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
! Calcium	Diasys 200 Pro	10.05	mg/dl	2.65
! HDL	Diasys 200 Pro	20.00	mg/dl	-2.16
Legend @ : z score $\leq$ 2.0 - Acceptable ! : $2.0 < z \text{ score} < 3.0$ - Warning X : z score $\geq 3$ - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Diasys 200 Pro


Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	42.30	40.67	1.13	-0.43
@ Creatinine	mg/dl	0.70	0.85	-1.34	-0.43
@ Bilirubin Total	mg/dl	0.50	0.45	0.68	0.54
@ Bilirubin Direct	mg/dl	0.30	0.22	0.70	0.64
@ Total Protein	g/dL	5.98	6.81	-1.21	-1.05
@ Albumin	g/dL	3.90	4.20	-0.74	-0.53
@ Uric Acid	mg/dl	4.77	4.36	0.71	-0.29
! Calcium	mg/dl	10.05	8.82	2.65	0.43
@ Cholesterol	mg/dl	173.00	170.36	0.04	0.15
@ Triglycerides	mg/dl	148.00	137.36	0.53	0.02
! HDL	mg/dl	20.00	31.25	-2.16	-0.36
@ LDL	mg/dl	102.03	99.01	0.07	0.13
@ ALT	U/L	17.40	21.41	-0.41	0.15
@ AST	U/L	38.75	41.16	0.20	1.11
@ ALP	U/L	83.40	70.28	0.15	-0.68

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	120.98	114.30	1.24	0.39
@ Sodium	mmol/L	158.63	162.40	-0.35	-0.32
@ Potassium	mmol/L	6.32	6.31	0.12	-0.09

Legend @ : Acceptable  
 ! : 2.0 < z score < 3.0 - Warning  
 X : z score ≥3 - Unacceptable  
 # : Not Evaluated  
 ⊙ : Delayed Result Entry  
 \* : Not considered for evaluation.

Total Parameters	18
Not Evaluated Parameters	0
Evaluated Parameters	18
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kirol Road, Kurla (W),  
 Mumbai - 400070