



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD DASUYA (1044)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 05

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_


**Dated:** \_\_\_\_\_

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Neutrophil.	-	40-70%	40-70%
@ Lymphocyte.	-	20-40%	20-40%
@ Eosinophil.	-	<6%	<6%
@ Monocyte.	-	1-10%	1-10%
@ Basophil.	-	<1%	<1%
@ Metamyelocyte.	-	Absent	Absent
@ Promyelocyte.	-	Absent	Absent
@ Myelocyte.	-	Absent	Absent

Legend @ : Acceptable  
 @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⊙ : Delayed Result Entry

Total Parameters	8
Not Evaluated Parameters	0
Evaluated Parameters	8
Outlier Parameters (X)	0
EQAS Score Haematology-PS Slide Count Qualitative	100.00 %



**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kirol Road, Kurla (W),  
 Mumbai - 400070



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD DASUYA (1044)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 05

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_


**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Diagnosis (Slide)	-	P. vivax	P. vivax
Legend @ : Acceptable		Total Parameters	1
@ : Acceptable		Not Evaluated Parameters	0
X : Unacceptable		Evaluated Parameters	1
# : Not Evaluated		Outlier Parameters (X)	0
⊙ : Delayed Result Entry		EQAS Score Haematology- Peripheral smear diagnosis	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kiroi Road, Kurla (W),  
Mumbai - 400070



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD DASUYA (1044)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 02

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.64	2.73	-0.33	-0.81
@ Total Leucocytes (WBC) count	/c.mm	7310.00	7767.57	-0.41	2.62
@ Haemoglobin	g/dL	8.30	8.06	-0.23	-0.16
@ MCV (Mean Corpuscular Volume)	fL	119.60	117.70	0.04	-0.65
@ MCH (Mean Corpuscular Hb)	pg	31.40	29.56	0.96	1.24
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	26.30	24.78	1.02	1.64
@ Platelet Count	1000/ $\mu$ L	239.00	271.83	-0.75	0.86
@ Hematocrit	%	31.60	31.93	-0.58	-1.11
@ RDW CV%	%	17.20	18.09	-0.43	0.02

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score  $\geq$ 3 - Unacceptable

# : Not Evaluated

⌚ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD DASUYA (1044)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.39	2.47	-0.30	-0.81
@ Total Leucocytes (WBC) count	/c.mm	2580.00	2649.14	-0.38	2.62
@ Haemoglobin	g/dL	6.70	6.75	-0.90	-0.16
@ MCV (Mean Corpuscular Volume)	fL	107.10	106.32	-0.12	-0.65
@ MCH (Mean Corpuscular Hb)	pg	28.20	27.25	0.98	1.24
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	26.30	25.58	0.44	1.64
@ Platelet Count	1000/ $\mu$ L	171.00	158.56	1.40	0.86
@ Hematocrit	%	25.60	26.17	-0.72	-1.11
@ RDW CV%	%	18.00	18.60	-0.11	0.02

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score  $\geq$ 3 - Unacceptable

# : Not Evaluated

⌚ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070