

## **CALIBRATION CERTIFICATE**

This is to certify that the Hematology Analyzer **YUMIZEN H550** bearing serial number: **210YADH04029** installed at **KRSNAA DIAGNOSTICS, DASUYA, PUNJAB** has calibrated on 16<sup>TH</sup> JANUARY 2024.

Calibrator : ABX MINOCAL  
Lot No. : CX490  
Expiry Date: 05/02/2024

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on **15th JANUARY 2025**.

A circular stamp with the text "HORIBA INDIA PVT. LTD." around the perimeter. Overlaid on the stamp is a handwritten signature in black ink.

Shrish Dixit

(Head- Products & Marketing)  
For **Horiba India Pvt. Ltd.**