



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,GURDASPUR (1039)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

| Analyte | Instrument | Result Value | Standard Unit | Z-Score |
|---|----------------|--------------|---------------|---------|
| X Uric Acid | Diasys 200 Pro | 2.43 | mg/dl | -4.86 |
| Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation. | | | | |

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro


| Analyte | Standard Unit | Result Value | Mean | Z-Score | RMZ |
|--------------------|---------------|--------------|--------|---------|-------|
| X Uric Acid | mg/dl | 2.43 | 4.36 | -4.86 | -1.16 |
| @ Urea | mg/dl | 39.20 | 40.67 | 0.54 | 1.08 |
| @ Creatinine | mg/dl | 0.60 | 0.85 | -1.91 | -0.42 |
| @ Bilirubin Total | mg/dl | 0.60 | 0.45 | 1.33 | 0.85 |
| @ Bilirubin Direct | mg/dl | 0.32 | 0.22 | 0.90 | 0.81 |
| @ Total Protein | g/dL | 6.87 | 6.81 | 0.44 | -0.50 |
| @ Albumin | g/dL | 4.00 | 4.20 | -0.37 | 1.14 |
| @ Calcium | mg/dl | 9.40 | 8.82 | 1.30 | 0.88 |
| @ Cholesterol | mg/dl | 166.00 | 170.36 | -0.51 | -0.68 |
| @ Triglycerides | mg/dl | 116.00 | 137.36 | -1.97 | -0.47 |
| @ HDL | mg/dl | 23.00 | 31.25 | -1.75 | -0.33 |
| @ LDL | mg/dl | 94.96 | 99.01 | -0.66 | -0.33 |
| @ ALT | U/L | 18.80 | 21.41 | -0.14 | 0.52 |
| @ AST | U/L | 30.90 | 41.16 | -0.67 | -0.63 |
| @ ALP | U/L | 60.84 | 70.28 | -1.40 | 0.10 |

Instrument : ACCULYTE-3P

| Analyte | Standard Unit | Result Value | Mean | Z-Score | RMZ |
|-------------|---------------|--------------|--------|---------|-------|
| @ Chloride | mmol/L | 110.64 | 114.30 | -0.26 | -0.71 |
| @ Sodium | mmol/L | 162.07 | 162.40 | -0.03 | -0.09 |
| @ Potassium | mmol/L | 6.45 | 6.31 | 0.34 | -0.19 |

Legend @ : Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score ≥3 - Unacceptable
: Not Evaluated
⊙ : Delayed Result Entry
* : Not considered for evaluation.

| | |
|--------------------------|---------|
| Total Parameters | 18 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 18 |
| Outlier Parameters (X) | 1 |
| EQAS Score Biochemistry | 94.44 % |



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070